

LABORATORY REPORT

If you have any questions concerning this report, please do not hesitate to call us at (800) 332-4345 or (574) 233-4777.

This report may not be reproduced, except in full, without written approval from EEA.

STATE CERTIFICATION LIST

State	Certification	State	Certification
Alabama	40700	Missouri	880
Alaska	IN00035	Montana	CERT0026
Arizona	AZ0432	Nebraska	NE-OS-05-04
Arkansas	IN00035	Nevada	IN00035
California	2920	New Hampshire*	2124
Colorado	IN035	New Jersey*	IN598
Colorado Radiochemistry	IN035	New Mexico	IN00035
Connecticut	PH-0132	New York*	11398
Delaware	IN035	North Carolina	18700
Florida*	E87775	North Dakota	R-035
Georgia	929	Ohio	87775
Hawaii	IN035	Oklahoma	D9508
Idaho	IN00035	Oregon (Primary AB)*	4074-001
Illinois*	200001	Pennsylvania*	68-00466
Illinois Microbiology	17767	Puerto Rico	IN00035
Illinois Radiochemistry	IN00035	Rhode Island	LAO00343
Indiana Chemistry	C-71-01	South Carolina	95005
Indiana Microbiology	M-76-07	South Dakota	IN00035
Iowa	098	Tennessee	TN02973
Kansas*	E-10233	Texas*	T104704187-15-8
Kentucky	90056	Texas/TCEQ	TX207
Louisiana*	LA180008	Utah*	IN00035
Maine	IN00035	Vermont	VT-8775
Maryland	209	Virginia*	460275
Massachusetts	M-IN035	Washington	C837
Michigan	9926	West Virginia	9927 C
Minnesota*	018-999-338	Wisconsin	999766900
Mississippi	IN035	Wyoming	IN035
EPA	IN00035		

*NELAP/TNI Recognized Accreditation Bodies

110 South Hill Street
 South Bend, IN 46617
 Tel: (574) 233-4777
 Fax: (574) 233-8207
 1 800 332 4345

Laboratory Report

Client: Cary/Apex WTP

Report: 409580

Attn: Rachel Monschein
 1400 Wimberly Road
 Apex, NC 27523

Priority: Standard Written

Status: Final

PWS ID: NC0392020

Sample Information					
EEA ID #	Client ID	Method	Collected Date / Time	Collected By:	Received Date / Time
3878307	Raw Water Intake (Wet Well)	L402	02/22/18 09:50	Client	02/23/18 09:30
3878308	Filter Effluent	L402	02/22/18 09:10	Client	02/23/18 09:30
3878309	Biofiltration Filter #3	L402	02/22/18 09:43	Client	02/23/18 09:30
3878310	Biofiltration Filter #4	L402	02/22/18 09:39	Client	02/23/18 09:30
3878311	Biofiltration Filter #5	L402	02/22/18 09:40	Client	02/23/18 09:30

Report Summary

Detailed quantitative results are presented on the following pages. The results presented relate only to the samples provided for analysis.

We appreciate the opportunity to provide you with this analysis. If you have any questions concerning this report, please do not hesitate to call Joseph Mattheis at (574) 233-4777.

Note: This report may not be reproduced, except in full, without written approval from EEA.

 Account Manager

Authorized Signature

Title

03/20/2018

Date

Client Name: Cary/Apex WTP

Report #: 409580

Sampling Point: Raw Water Intake (Wet Well)

PWS ID: NC0392020

EEA Methods									
Analyte ID #	Analyte	Method	Reg Limit	MRL†	Result	Units	Preparation Date	Analyzed Date	EEA ID #
120226-60-0	10:2 Fluorotelomer sulfonic acid (10:2 FTS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:00	3878307
757124-22-4	4:2 Fluorotelomer sulfonic acid (4:2 FTS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:00	3878307
27619-97-2	6:2 Fluorotelomer sulfonic acid (6:2 FTS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:00	3878307
39108-34-4	8:2 Fluorotelomer sulfonic acid (8:2 FTS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:00	3878307
958445-44-8	ADONA	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:00	3878307
73606-19-6	F-53B Major	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:00	3878307
83329-89-9	F-53B Minor	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:00	3878307
13252-13-6	GenX	L402	---	5.0	< 5.0	ng/L	03/08/18 08:17	03/09/18 03:00	3878307
4151-50-2	N-ethylperfluorooctane sulfonamide (NEtFOSA)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:00	3878307
1691-99-2	N-ethylperfluorooctane sulfonamidoethanol	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:00	3878307
31506-32-8	N-methylperfluorooctane sulfonamide (NMeFOSA)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:00	3878307
24448-09-7	N-methylperfluorooctane sulfonamidoethanol	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:00	3878307
375-73-5	Perfluorobutanesulfonic acid (PFBS)	L402	---	2.0	5.3	ng/L	03/08/18 08:17	03/09/18 03:00	3878307
375-22-4	Perfluorobutanoic acid (PFBA)	L402	---	5.0	22	ng/L	03/08/18 08:17	03/09/18 03:00	3878307
335-76-2	Perfluorodecanoic acid (PFDA)	L402	---	2.0	3.9	ng/L	03/08/18 08:17	03/09/18 03:00	3878307
375-85-9	Perfluoroheptanoic acid (PFHpA)	L402	---	2.0	36	ng/L	03/08/18 08:17	03/09/18 03:00	3878307
355-46-4	Perfluorohexanesulfonic acid (PFHxS)	L402	---	2.0	4.8	ng/L	03/08/18 08:17	03/09/18 03:00	3878307
307-24-4	Perfluorohexanoic acid (PFHxA)	L402	---	2.0	53	ng/L	03/08/18 08:17	03/09/18 03:00	3878307
307-55-1	Perfluorododecanoic acid (PFDoA)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:00	3878307
375-95-1	Perfluorononanoic acid (PFNA)	L402	---	2.0	4.4	ng/L	03/08/18 08:17	03/09/18 03:00	3878307
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	L402	---	2.0	16	ng/L	03/08/18 08:17	03/09/18 03:00	3878307
2991-50-6	N-ethyl Perfluorooctanesulfonamidoacetic acid	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:00	3878307
2355-31-9	N-methyl Perfluorooctanesulfonamidoacetic acid	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:00	3878307
335-67-1	Perfluorooctanoic acid (PFOA)	L402	---	2.0	19	ng/L	03/08/18 08:17	03/09/18 03:00	3878307
72629-94-8	Perfluorotridecanoic acid (PFTrDA)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:00	3878307
2058-94-8	Perfluoroundecanoic acid (PFUnA)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:00	3878307
NA	Perfluorododecanesulfonic acid (PFDoS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:00	3878307
335-77-3	Perfluorodecanesulfonic acid (PFDS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:00	3878307
375-92-8	Perfluoroheptanesulfonic acid (PFHpS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:00	3878307
67905-19-5	Perfluorohexadecanoic acid (PFHxDA)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:00	3878307
151772-58-6	Perfluoro-2-methoxyethoxyacetic acid	L402	---	5.0	< 5.0	ng/L	03/08/18 08:17	03/09/18 03:00	3878307
801212-59-9	Perfluoro-4-isopropoxybutanoic acid	L402	---	5.0	< 5.0	ng/L	03/08/18 08:17	03/09/18 03:00	3878307
863090-89-5	Perfluoro-4-methoxybutanoic acid (PFMOBA)	L402	---	5.0	< 5.0	ng/L	03/08/18 08:17	03/09/18 03:00	3878307
377-73-1	Perfluoro-3-methoxypropanoic acid (PFMOPrA)	L402	---	5.0	< 5.0	ng/L	03/08/18 08:17	03/09/18 03:00	3878307
68259-12-1	Perfluorononanesulfonic acid (PFNS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:00	3878307
754-91-6	Perfluorooctane sulfonamide (PFOSA)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:00	3878307
2706-90-3	Perfluoropentanoic acid (PFPeA)	L402	---	2.0	42	ng/L	03/08/18 08:17	03/09/18 03:00	3878307
2706-91-4	Perfluoropentanesulfonic acid (PFPeS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:00	3878307
376-06-7	Perfluorotetradecanoic acid (PFTeDA)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:00	3878307

Sampling Point: Filter Effluent

PWS ID: NC0392020

EEA Methods									
Analyte ID #	Analyte	Method	Reg Limit	MRL†	Result	Units	Preparation Date	Analyzed Date	EEA ID #
120226-60-0	10:2 Fluorotelomer sulfonic acid (10:2 FTS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:26	3878308
757124-22-4	4:2 Fluorotelomer sulfonic acid (4:2 FTS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:26	3878308
27619-97-2	6:2 Fluorotelomer sulfonic acid (6:2 FTS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:26	3878308
39108-34-4	8:2 Fluorotelomer sulfonic acid (8:2 FTS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:26	3878308
958445-44-8	ADONA	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:26	3878308
73606-19-6	F-53B Major	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:26	3878308
83329-89-9	F-53B Minor	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:26	3878308
13252-13-6	GenX	L402	---	5.0	< 5.0	ng/L	03/08/18 08:17	03/09/18 03:26	3878308
4151-50-2	N-ethylperfluorooctane sulfonamide (NEtFOSA)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:26	3878308
1691-99-2	N-ethylperfluorooctane sulfonamidoethanol	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:26	3878308
31506-32-8	N-methylperfluorooctane sulfonamide (NMeFOSA)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:26	3878308
24448-09-7	N-methylperfluorooctane sulfonamidoethanol	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:26	3878308
375-73-5	Perfluorobutanesulfonic acid (PFBS)	L402	---	2.0	2.6	ng/L	03/08/18 08:17	03/09/18 03:26	3878308
375-22-4	Perfluorobutanoic acid (PFBA)	L402	---	5.0	19	ng/L	03/08/18 08:17	03/09/18 03:26	3878308
335-76-2	Perfluorodecanoic acid (PFDA)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:26	3878308
375-85-9	Perfluoroheptanoic acid (PFHpA)	L402	---	2.0	15	ng/L	03/08/18 08:17	03/09/18 03:26	3878308
355-46-4	Perfluorohexanesulfonic acid (PFHxS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:26	3878308
307-24-4	Perfluorohexanoic acid (PFHxA)	L402	---	2.0	31	ng/L	03/08/18 08:17	03/09/18 03:26	3878308
307-55-1	Perfluorododecanoic acid (PFDoA)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:26	3878308
375-95-1	Perfluorononanoic acid (PFNA)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:26	3878308
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:26	3878308
2991-50-6	N-ethyl Perfluorooctanesulfonamidoacetic acid	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:26	3878308
2355-31-9	N-methyl Perfluorooctanesulfonamidoacetic acid	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:26	3878308
335-67-1	Perfluorooctanoic acid (PFOA)	L402	---	2.0	5.9	ng/L	03/08/18 08:17	03/09/18 03:26	3878308
72629-94-8	Perfluorotridecanoic acid (PFTrDA)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:26	3878308
2058-94-8	Perfluoroundecanoic acid (PFUnA)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:26	3878308
NA	Perfluorododecanesulfonic acid (PFDoS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:26	3878308
335-77-3	Perfluorodecanesulfonic acid (PFDS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:26	3878308
375-92-8	Perfluoroheptanesulfonic acid (PFHpS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:26	3878308
67905-19-5	Perfluorohexadecanoic acid (PFHxDA)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:26	3878308
151772-58-6	Perfluoro-2-methoxyethoxyacetic acid	L402	---	5.0	< 5.0	ng/L	03/08/18 08:17	03/09/18 03:26	3878308
801212-59-9	Perfluoro-4-isopropoxybutanoic acid	L402	---	5.0	< 5.0	ng/L	03/08/18 08:17	03/09/18 03:26	3878308
863090-89-5	Perfluoro-4-methoxybutanoic acid (PFMOBA)	L402	---	5.0	< 5.0	ng/L	03/08/18 08:17	03/09/18 03:26	3878308
377-73-1	Perfluoro-3-methoxypropanoic acid (PFMOPrA)	L402	---	5.0	< 5.0	ng/L	03/08/18 08:17	03/09/18 03:26	3878308
68259-12-1	Perfluorononanesulfonic acid (PFNS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:26	3878308
754-91-6	Perfluorooctane sulfonamide (PFOSA)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:26	3878308
2706-90-3	Perfluoropentanoic acid (PFPeA)	L402	---	2.0	33	ng/L	03/08/18 08:17	03/09/18 03:26	3878308
2706-91-4	Perfluoropentanesulfonic acid (PFPeS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:26	3878308
376-06-7	Perfluorotetradecanoic acid (PFTeDA)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:26	3878308

Sampling Point: Biofiltration Filter #3

PWS ID: NC0392020

EEA Methods									
Analyte ID #	Analyte	Method	Reg Limit	MRL†	Result	Units	Preparation Date	Analyzed Date	EEA ID #
120226-60-0	10:2 Fluorotelomer sulfonic acid (10:2 FTS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:52	3878309
757124-22-4	4:2 Fluorotelomer sulfonic acid (4:2 FTS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:52	3878309
27619-97-2	6:2 Fluorotelomer sulfonic acid (6:2 FTS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:52	3878309
39108-34-4	8:2 Fluorotelomer sulfonic acid (8:2 FTS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:52	3878309
958445-44-8	ADONA	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:52	3878309
73606-19-6	F-53B Major	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:52	3878309
83329-89-9	F-53B Minor	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:52	3878309
13252-13-6	GenX	L402	---	5.0	< 5.0	ng/L	03/08/18 08:17	03/09/18 03:52	3878309
4151-50-2	N-ethylperfluorooctane sulfonamide (NEtFOSA)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:52	3878309
1691-99-2	N-ethylperfluorooctane sulfonamidoethanol	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:52	3878309
31506-32-8	N-methylperfluorooctane sulfonamide (NMeFOSA)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:52	3878309
24448-09-7	N-methylperfluorooctane sulfonamidoethanol	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:52	3878309
375-73-5	Perfluorobutanesulfonic acid (PFBS)	L402	---	2.0	2.5	ng/L	03/08/18 08:17	03/09/18 03:52	3878309
375-22-4	Perfluorobutanoic acid (PFBA)	L402	---	5.0	19	ng/L	03/08/18 08:17	03/09/18 03:52	3878309
335-76-2	Perfluorodecanoic acid (PFDA)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:52	3878309
375-85-9	Perfluoroheptanoic acid (PFHpA)	L402	---	2.0	13	ng/L	03/08/18 08:17	03/09/18 03:52	3878309
355-46-4	Perfluorohexanesulfonic acid (PFHxS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:52	3878309
307-24-4	Perfluorohexanoic acid (PFHxA)	L402	---	2.0	31	ng/L	03/08/18 08:17	03/09/18 03:52	3878309
307-55-1	Perfluorododecanoic acid (PFDoA)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:52	3878309
375-95-1	Perfluorononanoic acid (PFNA)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:52	3878309
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:52	3878309
2991-50-6	N-ethyl Perfluorooctanesulfonamidoacetic acid	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:52	3878309
2355-31-9	N-methyl Perfluorooctanesulfonamidoacetic acid	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:52	3878309
335-67-1	Perfluorooctanoic acid (PFOA)	L402	---	2.0	5.2	ng/L	03/08/18 08:17	03/09/18 03:52	3878309
72629-94-8	Perfluorotridecanoic acid (PFTrDA)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:52	3878309
2058-94-8	Perfluoroundecanoic acid (PFUnA)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:52	3878309
NA	Perfluorododecanesulfonic acid (PFDoS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:52	3878309
335-77-3	Perfluorodecanesulfonic acid (PFDS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:52	3878309
375-92-8	Perfluoroheptanesulfonic acid (PFHpS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:52	3878309
67905-19-5	Perfluorohexadecanoic acid (PFHxDA)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:52	3878309
151772-58-6	Perfluoro-2-methoxyethoxyacetic acid	L402	---	5.0	< 5.0	ng/L	03/08/18 08:17	03/09/18 03:52	3878309
801212-59-9	Perfluoro-4-isopropoxybutanoic acid	L402	---	5.0	< 5.0	ng/L	03/08/18 08:17	03/09/18 03:52	3878309
863090-89-5	Perfluoro-4-methoxybutanoic acid (PFMOBA)	L402	---	5.0	< 5.0	ng/L	03/08/18 08:17	03/09/18 03:52	3878309
377-73-1	Perfluoro-3-methoxypropanoic acid (PFMOPrA)	L402	---	5.0	< 5.0	ng/L	03/08/18 08:17	03/09/18 03:52	3878309
68259-12-1	Perfluorononanesulfonic acid (PFNS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:52	3878309
754-91-6	Perfluorooctane sulfonamide (PFOSA)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:52	3878309
2706-90-3	Perfluoropentanoic acid (PFPeA)	L402	---	2.0	33	ng/L	03/08/18 08:17	03/09/18 03:52	3878309
2706-91-4	Perfluoropentanesulfonic acid (PFPeS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:52	3878309
376-06-7	Perfluorotetradecanoic acid (PFTeDA)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 03:52	3878309

Sampling Point: Biofiltration Filter #4

PWS ID: NC0392020

EEA Methods									
Analyte ID #	Analyte	Method	Reg Limit	MRL†	Result	Units	Preparation Date	Analyzed Date	EEA ID #
120226-60-0	10:2 Fluorotelomer sulfonic acid (10:2 FTS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:18	3878310
757124-22-4	4:2 Fluorotelomer sulfonic acid (4:2 FTS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:18	3878310
27619-97-2	6:2 Fluorotelomer sulfonic acid (6:2 FTS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:18	3878310
39108-34-4	8:2 Fluorotelomer sulfonic acid (8:2 FTS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:18	3878310
958445-44-8	ADONA	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:18	3878310
73606-19-6	F-53B Major	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:18	3878310
83329-89-9	F-53B Minor	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:18	3878310
13252-13-6	GenX	L402	---	5.0	< 5.0	ng/L	03/08/18 08:17	03/09/18 04:18	3878310
4151-50-2	N-ethylperfluorooctane sulfonamide (NEtFOSA)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:18	3878310
1691-99-2	N-ethylperfluorooctane sulfonamidoethanol	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:18	3878310
31506-32-8	N-methylperfluorooctane sulfonamide (NMeFOSA)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:18	3878310
24448-09-7	N-methylperfluorooctane sulfonamidoethanol	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:18	3878310
375-73-5	Perfluorobutanesulfonic acid (PFBS)	L402	---	2.0	2.4	ng/L	03/08/18 08:17	03/09/18 04:18	3878310
375-22-4	Perfluorobutanoic acid (PFBA)	L402	---	5.0	19	ng/L	03/08/18 08:17	03/09/18 04:18	3878310
335-76-2	Perfluorodecanoic acid (PFDA)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:18	3878310
375-85-9	Perfluoroheptanoic acid (PFHpA)	L402	---	2.0	12	ng/L	03/08/18 08:17	03/09/18 04:18	3878310
355-46-4	Perfluorohexanesulfonic acid (PFHxS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:18	3878310
307-24-4	Perfluorohexanoic acid (PFHxA)	L402	---	2.0	30	ng/L	03/08/18 08:17	03/09/18 04:18	3878310
307-55-1	Perfluorododecanoic acid (PFDoA)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:18	3878310
375-95-1	Perfluorononanoic acid (PFNA)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:18	3878310
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:18	3878310
2991-50-6	N-ethyl Perfluorooctanesulfonamidoacetic acid	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:18	3878310
2355-31-9	N-methyl Perfluorooctanesulfonamidoacetic acid	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:18	3878310
335-67-1	Perfluorooctanoic acid (PFOA)	L402	---	2.0	4.3	ng/L	03/08/18 08:17	03/09/18 04:18	3878310
72629-94-8	Perfluorotridecanoic acid (PFTrDA)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:18	3878310
2058-94-8	Perfluoroundecanoic acid (PFUnA)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:18	3878310
NA	Perfluorododecanesulfonic acid (PFDoS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:18	3878310
335-77-3	Perfluorodecanesulfonic acid (PFDS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:18	3878310
375-92-8	Perfluoroheptanesulfonic acid (PFHpS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:18	3878310
67905-19-5	Perfluorohexadecanoic acid (PFHxDA)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:18	3878310
151772-58-6	Perfluoro-2-methoxyethoxyacetic acid	L402	---	5.0	< 5.0	ng/L	03/08/18 08:17	03/09/18 04:18	3878310
801212-59-9	Perfluoro-4-isopropoxybutanoic acid	L402	---	5.0	< 5.0	ng/L	03/08/18 08:17	03/09/18 04:18	3878310
863090-89-5	Perfluoro-4-methoxybutanoic acid (PFMOBA)	L402	---	5.0	< 5.0	ng/L	03/08/18 08:17	03/09/18 04:18	3878310
377-73-1	Perfluoro-3-methoxypropanoic acid (PFMOPrA)	L402	---	5.0	< 5.0	ng/L	03/08/18 08:17	03/09/18 04:18	3878310
68259-12-1	Perfluorononanesulfonic acid (PFNS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:18	3878310
754-91-6	Perfluorooctane sulfonamide (PFOSA)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:18	3878310
2706-90-3	Perfluoropentanoic acid (PFPeA)	L402	---	2.0	32	ng/L	03/08/18 08:17	03/09/18 04:18	3878310
2706-91-4	Perfluoropentanesulfonic acid (PFPeS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:18	3878310
376-06-7	Perfluorotetradecanoic acid (PFTeDA)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:18	3878310

EEA Methods									
Analyte ID #	Analyte	Method	Reg Limit	MRL†	Result	Units	Preparation Date	Analyzed Date	EEA ID #
120226-60-0	10:2 Fluorotelomer sulfonic acid (10:2 FTS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:44	3878311
757124-22-4	4:2 Fluorotelomer sulfonic acid (4:2 FTS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:44	3878311
27619-97-2	6:2 Fluorotelomer sulfonic acid (6:2 FTS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:44	3878311
39108-34-4	8:2 Fluorotelomer sulfonic acid (8:2 FTS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:44	3878311
958445-44-8	ADONA	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:44	3878311
73606-19-6	F-53B Major	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:44	3878311
83329-89-9	F-53B Minor	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:44	3878311
13252-13-6	GenX	L402	---	5.0	< 5.0	ng/L	03/08/18 08:17	03/09/18 04:44	3878311
4151-50-2	N-ethylperfluorooctane sulfonamide (NEtFOSA)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:44	3878311
1691-99-2	N-ethylperfluorooctane sulfonamidoethanol	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:44	3878311
31506-32-8	N-methylperfluorooctane sulfonamide (NMeFOSA)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:44	3878311
24448-09-7	N-methylperfluorooctane sulfonamidoethanol	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:44	3878311
375-73-5	Perfluorobutanesulfonic acid (PFBS)	L402	---	2.0	2.9	ng/L	03/08/18 08:17	03/09/18 04:44	3878311
375-22-4	Perfluorobutanoic acid (PFBA)	L402	---	5.0	20	ng/L	03/08/18 08:17	03/09/18 04:44	3878311
335-76-2	Perfluorodecanoic acid (PFDA)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:44	3878311
375-85-9	Perfluoroheptanoic acid (PFHpA)	L402	---	2.0	16	ng/L	03/08/18 08:17	03/09/18 04:44	3878311
355-46-4	Perfluorohexanesulfonic acid (PFHxS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:44	3878311
307-24-4	Perfluorohexanoic acid (PFHxA)	L402	---	2.0	34	ng/L	03/08/18 08:17	03/09/18 04:44	3878311
307-55-1	Perfluorododecanoic acid (PFDoA)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:44	3878311
375-95-1	Perfluorononanoic acid (PFNA)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:44	3878311
1763-23-1	Perfluorooctanesulfonic acid (PFOS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:44	3878311
2991-50-6	N-ethyl Perfluorooctanesulfonamidoacetic acid	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:44	3878311
2355-31-9	N-methyl Perfluorooctanesulfonamidoacetic acid	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:44	3878311
335-67-1	Perfluorooctanoic acid (PFOA)	L402	---	2.0	6.5	ng/L	03/08/18 08:17	03/09/18 04:44	3878311
72629-94-8	Perfluorotridecanoic acid (PFTrDA)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:44	3878311
2058-94-8	Perfluoroundecanoic acid (PFUnA)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:44	3878311
NA	Perfluorododecanesulfonic acid (PFDoS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:44	3878311
335-77-3	Perfluorodecanesulfonic acid (PFDS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:44	3878311
375-92-8	Perfluoroheptanesulfonic acid (PFHpS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:44	3878311
67905-19-5	Perfluorohexadecanoic acid (PFHxDA)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:44	3878311
151772-58-6	Perfluoro-2-methoxyethoxyacetic acid	L402	---	5.0	< 5.0	ng/L	03/08/18 08:17	03/09/18 04:44	3878311
801212-59-9	Perfluoro-4-isopropoxybutanoic acid	L402	---	5.0	< 5.0	ng/L	03/08/18 08:17	03/09/18 04:44	3878311
863090-89-5	Perfluoro-4-methoxybutanoic acid (PFMOBA)	L402	---	5.0	< 5.0	ng/L	03/08/18 08:17	03/09/18 04:44	3878311
377-73-1	Perfluoro-3-methoxypropanoic acid (PFMOPrA)	L402	---	5.0	< 5.0	ng/L	03/08/18 08:17	03/09/18 04:44	3878311
68259-12-1	Perfluorononanesulfonic acid (PFNS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:44	3878311
754-91-6	Perfluorooctane sulfonamide (PFOSA)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:44	3878311
2706-90-3	Perfluoropentanoic acid (PFPeA)	L402	---	2.0	34	ng/L	03/08/18 08:17	03/09/18 04:44	3878311
2706-91-4	Perfluoropentanesulfonic acid (PFPeS)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:44	3878311
376-06-7	Perfluorotetradecanoic acid (PFTeDA)	L402	---	2.0	< 2.0	ng/L	03/08/18 08:17	03/09/18 04:44	3878311

† EEA has demonstrated it can achieve these report limits in reagent water, but can not document them in all sample matrices.

Reg Limit Type:	MCL	SMCL	AL
Symbol:	*	^	!

Lab Definitions

Continuing Calibration Check Standard (CCC) / Continuing Calibration Verification (CCV) / Initial Calibration Verification Standard (ICV) / Initial Performance Check (IPC) - is a standard containing one or more of the target analytes that is prepared from the same standards used to calibrate the instrument. This standard is used to verify the calibration curve at the beginning of each analytical sequence, and may also be analyzed throughout and at the end of the sequence. The concentration of continuing standards may be varied, when prescribed by the reference method, so that the range of the calibration curve is verified on a regular basis. CCL, CCM, and CCH are the CCC standards at low, mid, and high concentration levels, respectively.

Internal Standards (IS) - are pure compounds with properties similar to the analytes of interest, which are added to field samples or extracts, calibration standards, and quality control standards at a known concentration. They are used to measure the relative responses of the analytes of interest and surrogates in the sample, calibration standard or quality control standard.

Laboratory Duplicate (LD) - is a field sample aliquot taken from the same sample container in the laboratory and analyzed separately using identical procedures. Analysis of laboratory duplicates provides a measure of the precision of the laboratory procedures.

Laboratory Fortified Blank (LFB) / Laboratory Control Sample (LCS) - is an aliquot of reagent water to which known concentrations of the analytes of interest are added. The LFB is analyzed exactly the same as the field samples. LFBs are used to determine whether the method is in control. FBL, FBM, and FBH are the LFB samples at low, mid, and high concentration levels, respectively.

Laboratory Method Blank (LMB) / Laboratory Reagent Blank (LRB) - is a sample of reagent water included in the sample batch analyzed in the same way as the associated field samples. The LMB is used to determine if method analytes or other background contamination have been introduced during the preparation or analytical procedure. The LMB is analyzed exactly the same as the field samples.

Laboratory Trip Blank (LTB) / Field Reagent Blank (FRB) - is a sample of laboratory reagent water placed in a sample container in the laboratory and treated as a field sample, including storage, preservation, and all analytical procedures. The FRB/LTB container follows the collection bottles to and from the collection site, but the FRB/LTB is not opened at any time during the trip. The FRB/LTB is primarily a travel blank used to verify that the samples were not contaminated during shipment.

Matrix Spike Duplicate Sample (MSD) / Laboratory Fortified Sample Matrix Duplicate (LFSMD) - is a sample aliquot taken from the same field sample source as the Matrix Spike Sample to which known quantities of the analytes of interest are added in the laboratory. The MSD is analyzed exactly the same as the field samples. Analysis of the MSD provides a measure of the precision of the laboratory procedures in a specific matrix. SDL, SDM, and SDH / LFSMDL, LFSMDM, and LFSMDH are the MSD or LFSMD at low, mid, and high concentration levels, respectively.

Matrix Spike Sample (MS) / Laboratory Fortified Sample Matrix (LFSM) - is a sample aliquot taken from field sample source to which known quantities of the analytes of interest are added in the laboratory. The MS is analyzed exactly the same as the field samples. The purpose is to demonstrate recovery of the analytes from a sample matrix to determine if the specific matrix contributes bias to the analytical results. MSL, MSM, and MSH / LFSML, LFSMM, and LFSMH are the MS or LFSM at low, mid, and high concentration levels, respectively.

Quality Control Standard (QCS) / Second Source Calibration Verification (SSCV) - is a solution containing known concentrations of the analytes of interest prepared from a source different from the source of the calibration standards. The solution is obtained from a second manufacturer or lot if the lot can be demonstrated by the manufacturer as prepared independently from other lots. The QCS sample is analyzed using the same procedures as field samples. The QCS is used as a check on the calibration standards used in the method on a routine basis.

Reporting Limit Check (RLC) / Initial Calibration Check Standard (ICCS) - is a procedural standard that is analyzed each day to evaluate instrument performance at or below the minimum reporting limit (MRL).

Surrogate Standard (SS) / Surrogate Analyte (SUR) - is a pure compound with properties similar to the analytes of interest, which is highly unlikely to be found in any field sample, that is added to the field samples, calibration standards, blanks and quality control standards before sample preparation. The SS is used to evaluate the efficiency of the sample preparation process.



Eaton Analytical

110 S. Hill Street
South Bend, IN 46617
T: 1.800.332.4345
F: 1.574.233.8207

Order # 335428
Batch # 409580

www.EurofinsUS.com/Eaton

CHAIN OF CUSTODY RECORD

Page 1 of 1

Shaded area for EEA use only

REPORT TO:	SAMPLER (Signature)	PWS ID #	STATE (sample origin)	PROJECT NAME	PO#	# OF CONTAINERS	MATRIX CODE	TURNAROUND TIME	COLLECTION		SAMPLING SITE	TEST NAME	SAMPLE REMARKS	CHLORINATED	
									DATE	TIME				YES	NO
Rachel Monschey 1400 Wimberville Rd Cary, NC 27523 Town of Cary Finance .AP. P.O. Box 8003 Cary, NC 27512-8005	<i>[Signature]</i> COMPLIANCE MONITORING	NC0392020	NC	PFAS		2	DW5W		2/22/18	9:50	Raw Water Intake Wetwell Filter equipment Biofiltration Filter #3 #4 #5	PH 2-3 5.2-5.3	X		
									2/22/18	9:10			X		
									2/22/18	9:43			X		
									2/22/18	9:39			X		
									2/22/18	9:40			X		
6															
7															
8															
9															
10															
11															
12															
13															
14															

RELINQUISHED BY: (Signature)	DATE	TIME	RECEIVED BY: (Signature)	DATE	TIME	LAB COMMENTS
<i>[Signature]</i>	2/22/18	10:15	<i>[Signature]</i>	02/22/18	3:30	
<i>[Signature]</i>	2/22/18	5:00	<i>[Signature]</i>			
<i>[Signature]</i>			<i>[Signature]</i>	2/23/18	09:30	

CONDITIONS UPON RECEIPT (check one):
 Ice/ Wet/ Blue Ambient °C Upon Receipt N/A

2.4

MATRIX CODES:
 DW-DRINKING WATER
 RW-REAGENT WATER
 GW-GROUND WATER
 EW-EXPOSURE WATER
 SW-SURFACE WATER
 PW-POOL WATER
 WW-WASTE WATER

TURN-AROUND TIME (TAT) - SURCHARGES

SW = Standard Written: (15 working days) 0%
 RV* = Rush Verbal: (5 working days) 50%
 RW* = Rush Written: (5 working days) 75%

IV* = Immediate Verbal: (3 working days) 100%
IW* = Immediate Written: (3 working days) 125%
SP* = Weekend, Holiday CALL
STAT* = Less than 48 hours CALL

*** Please call, expedited service not available for all testing**

06-LO-F0435 Issue 6.0 Effective Date: 2016-09-20



Eurofins Eaton Analytical

Run Log

Run ID: 240737 Method: L402

Type	Sample Id	Sample Site	Matrix	Instrument ID	Analysis Date	Calibration File
CCL	3886491		OS	FL	03/08/2018 20:55	030818L402a-PFCQuant-NTA.mdb
LRB	3886473		RW	FL	03/08/2018 21:21	030818L402a-PFCQuant-NTA.mdb
FBL	3886482		RW	FL	03/08/2018 21:47	030818L402a-PFCQuant-NTA.mdb
CCM	3886492		OS	FL	03/09/2018 02:34	030818L402a-PFCQuant-NTA.mdb
FS	3878307	Raw Water Intake (Wet Well)	DW	FL	03/09/2018 03:00	030818L402a-PFCQuant-NTA.mdb
FS	3878308	Filter Effluent	DW	FL	03/09/2018 03:26	030818L402a-PFCQuant-NTA.mdb
FS	3878309	Biofiltration Filter #3	DW	FL	03/09/2018 03:52	030818L402a-PFCQuant-NTA.mdb
FS	3878310	Biofiltration Filter #4	DW	FL	03/09/2018 04:18	030818L402a-PFCQuant-NTA.mdb
FS	3878311	Biofiltration Filter #5	DW	FL	03/09/2018 04:44	030818L402a-PFCQuant-NTA.mdb
CCH	3886493		OS	FL	03/09/2018 06:54	030818L402a-PFCQuant-NTA.mdb

QC Summary Report

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	DII Factor	Extracted	Analyzed	EEA ID #
CCL	IS-4-2-FTS-13C2	L402	N/A	---		429567	520489	ng/L	83	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	IS-6-2-FTS-13C2	L402	N/A	---		80808	82370	ng/L	98	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	IS-8-2-FTS-13C2	L402	N/A	---		403502	400513	ng/L	101	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	IS-GenX-13C3	L402	N/A	---		5906	6129	ng/L	96	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	IS-NEFOSAA-d5	L402	N/A	---		579809	552142	ng/L	105	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	IS-NEFOSE-d9	L402	N/A	---		81240	96472	ng/L	84	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	IS-NMeFOSAA-d3	L402	N/A	---		719591	696176	ng/L	103	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	IS-NMeFOSA-d3	L402	N/A	---		503812	588018	ng/L	86	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	IS-NMeFOSE-d7	L402	N/A	---		86872	100564	ng/L	86	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	IS-PFBA-13C4	L402	N/A	---		1710280	1971023	ng/L	87	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	IS-PFBS-13C3	L402	N/A	---		448911	497347	ng/L	90	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	IS-PFDA-13C2	L402	N/A	---		3529500	3124910	ng/L	113	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	IS-PFDoA-13C2	L402	N/A	---		2889170	2539471	ng/L	106	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	IS-PFHpa-13C4	L402	N/A	---		3441560	3493018	ng/L	99	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	IS-PFHXA-13C2	L402	N/A	---		3842920	4133503	ng/L	93	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	IS-PFOSA-13C8	L402	N/A	---		1126090	1118883	ng/L	101	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	IS-NEFOSA-d5	L402	N/A	---		440288	524732	ng/L	84	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	IS-PFHXS-13C3	L402	N/A	---		468956	498313	ng/L	94	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	IS-PFNA-13C9	L402	N/A	---		3477260	3311238	ng/L	105	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	IS-PFOA-13C2	L402	N/A	---		3855100	3737396	ng/L	103	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	IS-PFOS-13C4	L402	N/A	---		462409	443473	ng/L	104	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	IS-PFPeA-13C5	L402	N/A	---		1185980	1366858	ng/L	87	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	IS-PFTeDA-13C2	L402	N/A	---		1009170	1588255	ng/L	64	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	IS-PFUaA-13C7	L402	N/A	---		2968480	2529015	ng/L	117	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	10:2 Fluorotelomer sulfonic acid (10:2 FTS)	L402	2.0	---		2.1286	2.0	ng/L	106	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	4:2 Fluorotelomer sulfonic acid (4:2 FTS)	L402	2.0	---		2.1314	2.0	ng/L	107	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	6:2 Fluorotelomer sulfonic acid (6:2 FTS)	L402	2.0	---		1.8095	2.0	ng/L	90	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	8:2 Fluorotelomer sulfonic acid (8:2 FTS)	L402	2.0	---		2.0869	2.0	ng/L	104	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	ADONA	L402	2.0	---		1.8850	2.0	ng/L	94	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	F-53B Major	L402	2.0	---		1.4145	2.0	ng/L	71	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	F-53B Minor	L402	2.0	---		1.9263	2.0	ng/L	96	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	GenX	L402	5.0	---		4.4903	5.0	ng/L	90	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	N-ethylperfluorooctane sulfonamide (NEFOSA)	L402	2.0	---		1.9007	2.0	ng/L	95	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	N-ethylperfluorooctane sulfonamidoethanol	L402	2.0	---		1.8654	2.0	ng/L	93	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	N-methylperfluorooctane sulfonamide (NMeFOSA)	L402	2.0	---		1.8632	2.0	ng/L	93	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	N-methylperfluorooctane sulfonamidoethanol	L402	2.0	---		2.1763	2.0	ng/L	109	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	Perfluorobutanesulfonic acid (PFBS)	L402	2.0	---		1.7602	2.0	ng/L	88	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	Perfluorobutanoic acid (PFBA)	L402	5.0	---		4.3572	5.0	ng/L	87	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	Perfluorodecanoic acid (PFDA)	L402	2.0	---		1.7834	2.0	ng/L	89	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	Perfluoroheptanoic acid (PFHpA)	L402	2.0	---		1.7781	2.0	ng/L	89	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCL	Perfluorohexanesulfonic acid (PFHxS)	L402	2.0	---		1.7552	2.0	ng/L	88	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	Perfluorohexanoic acid (PFHxA)	L402	2.0	---		1.8347	2.0	ng/L	92	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	Perfluorodecanecarboxylic acid (PFDoA)	L402	2.0	---		1.8378	2.0	ng/L	92	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	Perfluorononanoic acid (PFNA)	L402	2.0	---		1.8087	2.0	ng/L	90	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	Perfluorooctanesulfonic acid (PFOS)	L402	2.0	---		1.7466	2.0	ng/L	87	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	N-ethyl Perfluorooctanesulfonamidoacetic acid	L402	2.0	---		1.9578	2.0	ng/L	98	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	N-methyl Perfluorooctanesulfonamidoacetic acid	L402	2.0	---		1.8412	2.0	ng/L	92	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	Perfluorodecanecarboxylic acid (PFDoA)	L402	2.0	---		1.7997	2.0	ng/L	90	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	Perfluorotridecanecarboxylic acid (PFTrDA)	L402	2.0	---		1.6341	2.0	ng/L	82	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	Perfluoroundecanoic acid (PFUnA)	L402	2.0	---		1.8085	2.0	ng/L	90	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	Perfluorododecanesulfonic acid (PFDoS)	L402	2.0	---		1.4417	2.0	ng/L	72	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	Perfluorodecanesulfonic acid (PFDS)	L402	2.0	---		1.7489	2.0	ng/L	87	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	Perfluoroheptanesulfonic acid (PFHpS)	L402	2.0	---		1.9166	2.0	ng/L	96	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	Perfluorohexadecanoic acid (PFHxDA)	L402	2.0	---		2.4121	2.0	ng/L	121	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	Perfluoro-2-methoxyethoxyacetic acid	L402	5.0	---		4.7294	5.0	ng/L	95	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	Perfluoro-4-isopropoxybutanoic acid	L402	5.0	---		4.9939	5.0	ng/L	100	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	Perfluoro-4-methoxybutanoic acid (PFMOBA)	L402	5.0	---		4.3975	5.0	ng/L	88	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	Perfluoro-3-methoxypropanoic acid (PFMOPRA)	L402	5.0	---		3.9603	5.0	ng/L	79	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	Perfluoronanesulfonic acid (PFNS)	L402	2.0	---		1.7273	2.0	ng/L	86	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	Perfluorooctane sulfonamide (PFOSA)	L402	2.0	---		1.7869	2.0	ng/L	89	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	Perfluoropentanoic acid (PFPeA)	L402	2.0	---		1.8528	2.0	ng/L	93	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	Perfluorotetradecanoic acid (PFTeDA)	L402	2.0	---		1.7467	2.0	ng/L	87	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
CCL	Perfluorotetradecanoic acid (PFTeDA)	L402	2.0	---		1.8704	2.0	ng/L	94	50 - 150	---	---	1.0	03/08/2018 14:29	03/08/2018 20:55	3886491
LRB	10:2 Fluorotelomer sulfonic acid (10:2 FTS)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.94	03/08/2018 08:17	03/08/2018 21:21	3886473
LRB	4:2 Fluorotelomer sulfonic acid (4:2 FTS)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.94	03/08/2018 08:17	03/08/2018 21:21	3886473
LRB	6:2 Fluorotelomer sulfonic acid (6:2 FTS)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.94	03/08/2018 08:17	03/08/2018 21:21	3886473
LRB	8:2 Fluorotelomer sulfonic acid (8:2 FTS)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.94	03/08/2018 08:17	03/08/2018 21:21	3886473
LRB	ADONA	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.94	03/08/2018 08:17	03/08/2018 21:21	3886473
LRB	F-53B Major	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.94	03/08/2018 08:17	03/08/2018 21:21	3886473
LRB	F-53B Minor	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.94	03/08/2018 08:17	03/08/2018 21:21	3886473
LRB	GenX	L402	5.0	---	<	5.0		ng/L	---	---	---	---	0.94	03/08/2018 08:17	03/08/2018 21:21	3886473
LRB	N-ethylperfluorooctane sulfonamide (NEFOSA)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.94	03/08/2018 08:17	03/08/2018 21:21	3886473
LRB	N-ethylperfluorooctane sulfonamidoethanol	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.94	03/08/2018 08:17	03/08/2018 21:21	3886473
LRB	N-methylperfluorooctane sulfonamide (NMeFOSA)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.94	03/08/2018 08:17	03/08/2018 21:21	3886473
LRB	N-methylperfluorooctane sulfonamidoethanol	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.94	03/08/2018 08:17	03/08/2018 21:21	3886473
LRB	Perfluorobutanesulfonic acid (PFBS)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.94	03/08/2018 08:17	03/08/2018 21:21	3886473
LRB	Perfluorobutanoic acid (PFBA)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.94	03/08/2018 08:17	03/08/2018 21:21	3886473
LRB	Perfluorodecanecarboxylic acid (PFDA)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.94	03/08/2018 08:17	03/08/2018 21:21	3886473
LRB	Perfluoroheptanoic acid (PFHpA)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.94	03/08/2018 08:17	03/08/2018 21:21	3886473
LRB	Perfluorohexanesulfonic acid (PFHxS)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.94	03/08/2018 08:17	03/08/2018 21:21	3886473
LRB	Perfluorohexanoic acid (PFHxA)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.94	03/08/2018 08:17	03/08/2018 21:21	3886473

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
LRB	Perfluorodecanoic acid (PFDoA)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.94	03/08/2018 08:17	03/08/2018 21:21	3886473
LRB	Perfluorononanoic acid (PFNA)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.94	03/08/2018 08:17	03/08/2018 21:21	3886473
LRB	Perfluorooctanesulfonic acid (PFOS)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.94	03/08/2018 08:17	03/08/2018 21:21	3886473
LRB	N-ethyl Perfluorooctanesulfonamidoacetic acid	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.94	03/08/2018 08:17	03/08/2018 21:21	3886473
LRB	N-methyl Perfluorooctanesulfonamidoacetic acid	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.94	03/08/2018 08:17	03/08/2018 21:21	3886473
LRB	Perfluorooctanoic acid (PFOA)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.94	03/08/2018 08:17	03/08/2018 21:21	3886473
LRB	Perfluorotridecanoic acid (PFTDA)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.94	03/08/2018 08:17	03/08/2018 21:21	3886473
LRB	Perfluoroundecanoic acid (PFUnA)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.94	03/08/2018 08:17	03/08/2018 21:21	3886473
LRB	Perfluorododecanesulfonic acid (PFDoS)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.94	03/08/2018 08:17	03/08/2018 21:21	3886473
LRB	Perfluorodecanesulfonic acid (PFDS)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.94	03/08/2018 08:17	03/08/2018 21:21	3886473
LRB	Perfluorohexadecanoic acid (PFHxS)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.94	03/08/2018 08:17	03/08/2018 21:21	3886473
LRB	Perfluorohexadecanoic acid (PFHxDA)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.94	03/08/2018 08:17	03/08/2018 21:21	3886473
LRB	Perfluoro-2-methoxyethoxyacetic acid	L402	5.0	---	<	5.0		ng/L	---	---	---	---	0.94	03/08/2018 08:17	03/08/2018 21:21	3886473
LRB	Perfluoro-4-isopropoxybutanoic acid	L402	5.0	---	<	5.0		ng/L	---	---	---	---	0.94	03/08/2018 08:17	03/08/2018 21:21	3886473
LRB	Perfluoro-4-methoxybutanoic acid (PFMOBA)	L402	5.0	---	<	5.0		ng/L	---	---	---	---	0.94	03/08/2018 08:17	03/08/2018 21:21	3886473
LRB	Perfluoro-3-methoxypropanoic acid (PFMOPRA)	L402	5.0	---	<	5.0		ng/L	---	---	---	---	0.94	03/08/2018 08:17	03/08/2018 21:21	3886473
LRB	Perfluoronanesulfonic acid (PFNS)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.94	03/08/2018 08:17	03/08/2018 21:21	3886473
LRB	Perfluorooctane sulfonamide (PFOSA)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.94	03/08/2018 08:17	03/08/2018 21:21	3886473
LRB	Perfluoropentanoic acid (PFPeA)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.94	03/08/2018 08:17	03/08/2018 21:21	3886473
LRB	Perfluoropentanesulfonic acid (PFPeS)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.94	03/08/2018 08:17	03/08/2018 21:21	3886473
LRB	Perfluorotetradecanoic acid (PFTeDA)	L402	2.0	---	<	2.0		ng/L	---	---	---	---	0.94	03/08/2018 08:17	03/08/2018 21:21	3886473
FBL	10:2 Fluorotelomer sulfonic acid (10:2 FTS)	L402	2.0	---	<	2.0982	2.0	ng/L	105	50 - 150	---	---	1.0	03/08/2018 08:17	03/08/2018 21:47	3886482
FBL	4:2 Fluorotelomer sulfonic acid (4:2 FTS)	L402	2.0	---	<	2.2977	2.0	ng/L	115	50 - 150	---	---	1.0	03/08/2018 08:17	03/08/2018 21:47	3886482
FBL	6:2 Fluorotelomer sulfonic acid (6:2 FTS)	L402	2.0	---	<	2.4360	2.0	ng/L	122	50 - 150	---	---	1.0	03/08/2018 08:17	03/08/2018 21:47	3886482
FBL	8:2 Fluorotelomer sulfonic acid (8:2 FTS)	L402	2.0	---	<	2.3777	2.0	ng/L	119	50 - 150	---	---	1.0	03/08/2018 08:17	03/08/2018 21:47	3886482
FBL	ADONA	L402	2.0	---	<	2.2302	2.0	ng/L	112	50 - 150	---	---	1.0	03/08/2018 08:17	03/08/2018 21:47	3886482
FBL	F-53B Major	L402	2.0	---	<	1.6210	2.0	ng/L	81	50 - 150	---	---	1.0	03/08/2018 08:17	03/08/2018 21:47	3886482
FBL	F-53B Minor	L402	2.0	---	<	2.1674	2.0	ng/L	108	50 - 150	---	---	1.0	03/08/2018 08:17	03/08/2018 21:47	3886482
FBL	GenX	L402	5.0	---	<	4.7452	5.0	ng/L	95	50 - 150	---	---	1.0	03/08/2018 08:17	03/08/2018 21:47	3886482
FBL	N-ethylperfluorooctane sulfonamide (NEFOSA)	L402	2.0	---	<	2.2825	2.0	ng/L	114	50 - 150	---	---	1.0	03/08/2018 08:17	03/08/2018 21:47	3886482
FBL	N-ethylperfluorooctane sulfonamidoethanol	L402	2.0	---	<	2.2461	2.0	ng/L	112	50 - 150	---	---	1.0	03/08/2018 08:17	03/08/2018 21:47	3886482
FBL	N-methylperfluorooctane sulfonamide (NMeFOSA)	L402	2.0	---	<	2.1082	2.0	ng/L	105	50 - 150	---	---	1.0	03/08/2018 08:17	03/08/2018 21:47	3886482
FBL	N-methylperfluorooctane sulfonamidoethanol	L402	2.0	---	<	2.5821	2.0	ng/L	129	50 - 150	---	---	1.0	03/08/2018 08:17	03/08/2018 21:47	3886482
FBL	Perfluorobutanesulfonic acid (PFBS)	L402	2.0	---	<	1.8947	2.0	ng/L	95	50 - 150	---	---	1.0	03/08/2018 08:17	03/08/2018 21:47	3886482
FBL	Perfluorobutanoic acid (PFBA)	L402	5.0	---	<	4.9680	5.0	ng/L	99	50 - 150	---	---	1.0	03/08/2018 08:17	03/08/2018 21:47	3886482
FBL	Perfluorodecanoic acid (PFDA)	L402	2.0	---	<	2.0073	2.0	ng/L	100	50 - 150	---	---	1.0	03/08/2018 08:17	03/08/2018 21:47	3886482
FBL	Perfluoroheptanoic acid (PFHpA)	L402	2.0	---	<	1.9675	2.0	ng/L	98	50 - 150	---	---	1.0	03/08/2018 08:17	03/08/2018 21:47	3886482
FBL	Perfluorohexanesulfonic acid (PFHxS)	L402	2.0	---	<	1.9448	2.0	ng/L	97	50 - 150	---	---	1.0	03/08/2018 08:17	03/08/2018 21:47	3886482
FBL	Perfluorohexanoic acid (PFHxA)	L402	2.0	---	<	2.0560	2.0	ng/L	103	50 - 150	---	---	1.0	03/08/2018 08:17	03/08/2018 21:47	3886482
FBL	Perfluorodecanoic acid (PFDoA)	L402	2.0	---	<	2.0332	2.0	ng/L	102	50 - 150	---	---	1.0	03/08/2018 08:17	03/08/2018 21:47	3886482
FBL	Perfluorononanoic acid (PFNA)	L402	2.0	---	<	2.0447	2.0	ng/L	102	50 - 150	---	---	1.0	03/08/2018 08:17	03/08/2018 21:47	3886482

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
FBL	Perfluorooctanesulfonic acid (PFOS)	L402	2.0	---		1.9886	2.0	ng/L	99	50 - 150	---	---	1.0	03/08/2018 08:17	03/08/2018 21:47	3886482
FBL	N-ethyl Perfluorooctanesulfonamidoacetic acid	L402	2.0	---		2.1757	2.0	ng/L	109	50 - 150	---	---	1.0	03/08/2018 08:17	03/08/2018 21:47	3886482
FBL	N-methyl Perfluorooctanesulfonamidoacetic acid	L402	2.0	---		2.0177	2.0	ng/L	101	50 - 150	---	---	1.0	03/08/2018 08:17	03/08/2018 21:47	3886482
FBL	Perfluorodecane sulfonic acid (PFDA)	L402	2.0	---		2.5717	2.0	ng/L	129	50 - 150	---	---	1.0	03/08/2018 08:17	03/08/2018 21:47	3886482
FBL	Perfluorododecane sulfonic acid (PFDoS)	L402	2.0	---		1.0151	2.0	ng/L	51	40 - 160	---	---	1.0	03/08/2018 08:17	03/08/2018 21:47	3886482
FBL	Perfluoroundecane sulfonic acid (PFUnA)	L402	2.0	---		1.9899	2.0	ng/L	99	50 - 150	---	---	1.0	03/08/2018 08:17	03/08/2018 21:47	3886482
FBL	Perfluorododecane sulfonic acid (PFDoS)	L402	2.0	---		0.8466	2.0	ng/L	42	40 - 160	---	---	1.0	03/08/2018 08:17	03/08/2018 21:47	3886482
FBL	Perfluorodecane sulfonic acid (PFDS)	L402	2.0	---		1.8342	2.0	ng/L	92	50 - 150	---	---	1.0	03/08/2018 08:17	03/08/2018 21:47	3886482
FBL	Perfluoroheptane sulfonic acid (PFHpS)	L402	2.0	---		2.2190	2.0	ng/L	111	50 - 150	---	---	1.0	03/08/2018 08:17	03/08/2018 21:47	3886482
FBL	Perfluorohexadecane sulfonic acid (PFHxDA)	L402	2.0	---		2.4110	2.0	ng/L	121	50 - 150	---	---	1.0	03/08/2018 08:17	03/08/2018 21:47	3886482
FBL	Perfluoro-2-methoxyethoxyacetic acid	L402	5.0	---		5.2823	5.0	ng/L	106	50 - 150	---	---	1.0	03/08/2018 08:17	03/08/2018 21:47	3886482
FBL	Perfluoro-4-isopropoxybutanoic acid	L402	5.0	---		5.5618	5.0	ng/L	111	50 - 150	---	---	1.0	03/08/2018 08:17	03/08/2018 21:47	3886482
FBL	Perfluoro-4-methoxybutanoic acid (PFMOBA)	L402	5.0	---		4.4575	5.0	ng/L	89	50 - 150	---	---	1.0	03/08/2018 08:17	03/08/2018 21:47	3886482
FBL	Perfluoro-3-methoxypropanoic acid (PFMOPRA)	L402	5.0	---		4.3912	5.0	ng/L	88	50 - 150	---	---	1.0	03/08/2018 08:17	03/08/2018 21:47	3886482
FBL	Perfluorononane sulfonic acid (PFNS)	L402	2.0	---		1.8042	2.0	ng/L	90	50 - 150	---	---	1.0	03/08/2018 08:17	03/08/2018 21:47	3886482
FBL	Perfluorooctane sulfonamide (PFOSA)	L402	2.0	---		1.9964	2.0	ng/L	100	50 - 150	---	---	1.0	03/08/2018 08:17	03/08/2018 21:47	3886482
FBL	Perfluoropentanoic acid (PFPeA)	L402	2.0	---		2.0432	2.0	ng/L	102	50 - 150	---	---	1.0	03/08/2018 08:17	03/08/2018 21:47	3886482
FBL	Perfluoropentane sulfonic acid (PFPeS)	L402	2.0	---		1.9184	2.0	ng/L	96	50 - 150	---	---	1.0	03/08/2018 08:17	03/08/2018 21:47	3886482
FBL	Perfluorotetradecane sulfonic acid (PFTeDA)	L402	2.0	---		2.0901	2.0	ng/L	105	50 - 150	---	---	1.0	03/08/2018 08:17	03/08/2018 21:47	3886482
CCM	IS-4-2-FTS-13C2	L402	N/A	---		260076	520489	ng/L	50	50 - 150	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	IS-6-2-FTS-13C2	L402	N/A	---		39081	82370	ng/L	47	50 - 150	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	IS-8-2-FTS-13C2	L402	N/A	---		166115	400513	ng/L	41	50 - 150	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	IS-GenX-13C3	L402	N/A	---		5876	6129	ng/L	96	50 - 150	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	IS-NEFOSAA-d5	L402	N/A	---		303067	552142	ng/L	55	50 - 150	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	IS-NEFOSE-d9	L402	N/A	---		63624	96472	ng/L	66	50 - 150	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	IS-NMeFOSAA-d3	L402	N/A	---		388283	696176	ng/L	56	50 - 150	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	IS-NMeFOSA-d3	L402	N/A	---		358780	588018	ng/L	61	50 - 150	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	IS-NMeFOSE-d7	L402	N/A	---		64788	100564	ng/L	64	50 - 150	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	IS-PFBA-13C4	L402	N/A	---		1478410	1971023	ng/L	75	50 - 150	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	IS-PFBS-13C3	L402	N/A	---		374543	497347	ng/L	75	50 - 150	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	IS-PFDA-13C2	L402	N/A	---		1925270	3124910	ng/L	62	50 - 150	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	IS-PFDoA-13C2	L402	N/A	---		1477670	2539471	ng/L	58	50 - 150	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	IS-PFHpa-13C4	L402	N/A	---		2475710	3493018	ng/L	71	50 - 150	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	IS-PFHXA-13C2	L402	N/A	---		3047070	4133503	ng/L	74	50 - 150	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	IS-PFOA-13C2	L402	N/A	---		650007	1118883	ng/L	58	50 - 150	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	IS-PFOA-13C8	L402	N/A	---		314283	524732	ng/L	60	50 - 150	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	IS-NEFOSA-d5	L402	N/A	---		364774	498313	ng/L	73	50 - 150	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	IS-PFHXS-13C3	L402	N/A	---		2188210	3311238	ng/L	66	50 - 150	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	IS-PFNA-13C9	L402	N/A	---		2575740	3737396	ng/L	69	50 - 150	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	IS-PFOA-13C2	L402	N/A	---		298899	443473	ng/L	67	50 - 150	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	IS-PFOA-13C4	L402	N/A	---		987994	1366858	ng/L	72	50 - 150	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCM	IS-PFTeA-13C2	L402	N/A	---		870934	1588255	ng/L	55	50 - 150	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	IS-PFUnA-13C7	L402	N/A	---		1393720	2529015	ng/L	55	50 - 150	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	10:2 Fluorotelomer sulfonic acid (10:2 FTS)	L402	2.0	---		77.5782	100	ng/L	78	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	4:2 Fluorotelomer sulfonic acid (4:2 FTS)	L402	2.0	---		108.9110	100	ng/L	109	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	6:2 Fluorotelomer sulfonic acid (6:2 FTS)	L402	2.0	---		100.0950	100	ng/L	100	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	8:2 Fluorotelomer sulfonic acid (8:2 FTS)	L402	2.0	---		109.1300	100	ng/L	109	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	ADONA	L402	2.0	---		107.1450	100	ng/L	107	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	F-53B Major	L402	2.0	---		99.3228	100	ng/L	99	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	F-53B Minor	L402	2.0	---		124.5820	100	ng/L	125	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	GenX	L402	5.0	---		242.9050	250	ng/L	97	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	N-ethylperfluorooctane sulfonamide (NEFOSA)	L402	2.0	---		101.8170	100	ng/L	102	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	N-ethylperfluorooctane sulfonamidoethanol	L402	2.0	---		99.5199	100	ng/L	100	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	N-methylperfluorooctane sulfonamide (NMeFOSA)	L402	2.0	---		100.0590	100	ng/L	100	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	N-methylperfluorooctane sulfonamidoethanol	L402	2.0	---		107.4040	100	ng/L	107	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	Perfluorobutanesulfonic acid (PFBS)	L402	2.0	---		98.6805	100	ng/L	99	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	Perfluorobutanoic acid (PFBA)	L402	5.0	---		249.4620	250	ng/L	100	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	Perfluorodecanoic acid (PFDA)	L402	2.0	---		99.8710	100	ng/L	100	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	Perfluoroheptanoic acid (PFHpA)	L402	2.0	---		98.7431	100	ng/L	99	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	Perfluorohexanesulfonic acid (PFHxS)	L402	2.0	---		100.5400	100	ng/L	101	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	Perfluorohexanoic acid (PFHxA)	L402	2.0	---		98.9679	100	ng/L	99	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	Perfluorododecanoic acid (PFDoA)	L402	2.0	---		101.9890	100	ng/L	102	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	Perfluorononanoic acid (PFNA)	L402	2.0	---		99.1550	100	ng/L	99	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	Perfluorooctanesulfonic acid (PFOS)	L402	2.0	---		99.7194	100	ng/L	100	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	N-ethyl Perfluorooctanesulfonamidoacetic acid	L402	2.0	---		101.5000	100	ng/L	102	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	N-methyl Perfluorooctanesulfonamidoacetic acid	L402	2.0	---		100.6410	100	ng/L	101	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	Perfluorooctanoic acid (PFOA)	L402	2.0	---		100.1210	100	ng/L	100	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	Perfluorotridecanoic acid (PFTrDA)	L402	2.0	---		93.1300	100	ng/L	93	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	Perfluoroundecanoic acid (PFUnA)	L402	2.0	---		101.8330	100	ng/L	102	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	Perfluorododecanesulfonic acid (PFDoS)	L402	2.0	---		104.0620	100	ng/L	104	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	Perfluorododecanesulfonic acid (PFDS)	L402	2.0	---		118.0490	100	ng/L	118	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	Perfluoroheptanesulfonic acid (PFHpS)	L402	2.0	---		104.5220	100	ng/L	105	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	Perfluorohexadecanoic acid (PFHxDA)	L402	2.0	---		100.6210	100	ng/L	101	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	Perfluoro-2-methoxyethoxyacetic acid	L402	5.0	---		257.6690	250	ng/L	103	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	Perfluoro-4-isopropoxybutanoic acid	L402	5.0	---		253.9210	250	ng/L	102	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	Perfluoro-4-methoxybutanoic acid (PFMOBA)	L402	5.0	---		244.9940	250	ng/L	98	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	Perfluoro-3-methoxypropanoic acid (PFMOPRA)	L402	5.0	---		253.5730	250	ng/L	101	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	Perfluorononanesulfonic acid (PFNS)	L402	2.0	---		99.7032	100	ng/L	100	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	Perfluorooctane sulfonamide (PFOSA)	L402	2.0	---		99.9364	100	ng/L	100	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	Perfluoropentanoic acid (PFPeA)	L402	2.0	---		100.2270	100	ng/L	100	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	Perfluoropentanesulfonic acid (PFPeS)	L402	2.0	---		98.9880	100	ng/L	99	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492
CCM	Perfluorotetradecanoic acid (PFTeDA)	L402	2.0	---		101.7460	100	ng/L	102	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 02:34	3886492

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	DII Factor	Extracted	Analyzed	EEA ID #
FS	10:2 Fluorotoluene sulfonic acid (10:2 FTS)	L402	2.0	Raw Water Intake (Wet Well)	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:00	3878307
FS	4:2 Fluorotoluene sulfonic acid (4:2 FTS)	L402	2.0	Raw Water Intake (Wet Well)	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:00	3878307
FS	6:2 Fluorotoluene sulfonic acid (6:2 FTS)	L402	2.0	Raw Water Intake (Wet Well)	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:00	3878307
FS	8:2 Fluorotoluene sulfonic acid (8:2 FTS)	L402	2.0	Raw Water Intake (Wet Well)	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:00	3878307
FS	ADONA	L402	2.0	Raw Water Intake (Wet Well)	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:00	3878307
FS	F-53B Major	L402	2.0	Raw Water Intake (Wet Well)	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:00	3878307
FS	F-53B Minor	L402	2.0	Raw Water Intake (Wet Well)	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:00	3878307
FS	GenX	L402	5.0	Raw Water Intake (Wet Well)	<	5.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:00	3878307
FS	N-ethylperfluorooctane sulfonamide (NEFOSA)	L402	2.0	Raw Water Intake (Wet Well)	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:00	3878307
FS	N-ethylperfluorooctane sulfonamidoethanol	L402	2.0	Raw Water Intake (Wet Well)	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:00	3878307
FS	N-methylperfluorooctane sulfonamide (NMeFOSA)	L402	2.0	Raw Water Intake (Wet Well)	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:00	3878307
FS	N-methylperfluorooctane sulfonamidoethanol	L402	2.0	Raw Water Intake (Wet Well)	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:00	3878307
FS	Perfluorobutanesulfonic acid (PFBS)	L402	2.0	Raw Water Intake (Wet Well)	<	5.3		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:00	3878307
FS	Perfluorobutanoic acid (PFBA)	L402	5.0	Raw Water Intake (Wet Well)	<	22		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:00	3878307
FS	Perfluorodecanoic acid (PFDA)	L402	2.0	Raw Water Intake (Wet Well)	<	3.9		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:00	3878307
FS	Perfluoroheptanoic acid (PFHpA)	L402	2.0	Raw Water Intake (Wet Well)	<	36		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:00	3878307
FS	Perfluorohexanesulfonic acid (PFHxS)	L402	2.0	Raw Water Intake (Wet Well)	<	4.8		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:00	3878307
FS	Perfluorohexanoic acid (PFHxA)	L402	2.0	Raw Water Intake (Wet Well)	<	53		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:00	3878307
FS	Perfluorododecanoic acid (PFDoA)	L402	2.0	Raw Water Intake (Wet Well)	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:00	3878307
FS	Perfluorooctanesulfonic acid (PFOS)	L402	2.0	Raw Water Intake (Wet Well)	<	4.4		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:00	3878307
FS	N-ethyl Perfluorooctanesulfonamidoacetic acid	L402	2.0	Raw Water Intake (Wet Well)	<	16		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:00	3878307
FS	N-methyl Perfluorooctanesulfonamidoacetic acid	L402	2.0	Raw Water Intake (Wet Well)	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:00	3878307
FS	Perfluorooctanoic acid (PFOA)	L402	2.0	Raw Water Intake (Wet Well)	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:00	3878307
FS	Perfluorodecane sulfonic acid (PFDS)	L402	2.0	Raw Water Intake (Wet Well)	<	19		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:00	3878307
FS	Perfluorotridecanoic acid (PFTDA)	L402	2.0	Raw Water Intake (Wet Well)	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:00	3878307
FS	Perfluoroundecanoic acid (PFUnA)	L402	2.0	Raw Water Intake (Wet Well)	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:00	3878307
FS	Perfluorododecane sulfonic acid (PFDoS)	L402	2.0	Raw Water Intake (Wet Well)	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:00	3878307
FS	Perfluorodecane sulfonic acid (PFDS)	L402	2.0	Raw Water Intake (Wet Well)	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:00	3878307
FS	Perfluoroheptanesulfonic acid (PFHpS)	L402	2.0	Raw Water Intake (Wet Well)	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:00	3878307
FS	Perfluorohexadecanoic acid (PFHxDA)	L402	2.0	Raw Water Intake (Wet Well)	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:00	3878307
FS	Perfluoro-2-methoxyethoxyacetic acid	L402	5.0	Raw Water Intake (Wet Well)	<	5.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:00	3878307
FS	Perfluoro-4-isopropoxybutanoic acid (PFMOBA)	L402	5.0	Raw Water Intake (Wet Well)	<	5.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:00	3878307
FS	Perfluoro-4-methoxybutanoic acid (PFMOBA)	L402	5.0	Raw Water Intake (Wet Well)	<	5.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:00	3878307
FS	Perfluoro-3-methoxypropanoic acid (PFMOPRA)	L402	5.0	Raw Water Intake (Wet Well)	<	5.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:00	3878307
FS	Perfluoronanesulfonic acid (PFNS)	L402	2.0	Raw Water Intake (Wet Well)	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:00	3878307
FS	Perfluorooctane sulfonamide (PFOSA)	L402	2.0	Raw Water Intake (Wet Well)	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:00	3878307
FS	Perfluoropentanoic acid (PFPeA)	L402	2.0	Raw Water Intake (Wet Well)	<	42		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:00	3878307
FS	Perfluoropentanesulfonic acid (PFPS)	L402	2.0	Raw Water Intake (Wet Well)	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:00	3878307
FS	Perfluorotetradecanoic acid (PFTeDA)	L402	2.0	Raw Water Intake (Wet Well)	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:00	3878307
FS	10:2 Fluorotoluene sulfonic acid (10:2 FTS)	L402	2.0	Filter Effluent	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:26	3878308
FS	4:2 Fluorotoluene sulfonic acid (4:2 FTS)	L402	2.0	Filter Effluent	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:26	3878308

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
FS	6:2 Fluorotelomer sulfonic acid (6:2 FTS)	L402	2.0	Filter Effluent	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:26	3878308
FS	8:2 Fluorotelomer sulfonic acid (8:2 FTS)	L402	2.0	Filter Effluent	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:26	3878308
FS	ADONA	L402	2.0	Filter Effluent	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:26	3878308
FS	F-53B Major	L402	2.0	Filter Effluent	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:26	3878308
FS	F-53B Minor	L402	2.0	Filter Effluent	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:26	3878308
FS	GenX	L402	5.0	Filter Effluent	<	5.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:26	3878308
FS	N-ethylperfluorooctane sulfonamide (NEFOSA)	L402	2.0	Filter Effluent	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:26	3878308
FS	N-ethylperfluorooctane sulfonamidoethanol	L402	2.0	Filter Effluent	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:26	3878308
FS	N-methylperfluorooctane sulfonamide (NMeFOSA)	L402	2.0	Filter Effluent	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:26	3878308
FS	N-methylperfluorooctane sulfonamidoethanol	L402	2.0	Filter Effluent	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:26	3878308
FS	Perfluorobutanesulfonic acid (PFBS)	L402	2.0	Filter Effluent	<	2.6		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:26	3878308
FS	Perfluorobutanoic acid (PFBA)	L402	5.0	Filter Effluent	<	19		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:26	3878308
FS	Perfluorodecanoic acid (PFDA)	L402	2.0	Filter Effluent	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:26	3878308
FS	Perfluoroheptanoic acid (PFHpA)	L402	2.0	Filter Effluent	<	15		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:26	3878308
FS	Perfluorohexanesulfonic acid (PFHxS)	L402	2.0	Filter Effluent	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:26	3878308
FS	Perfluorohexanoic acid (PFHxA)	L402	2.0	Filter Effluent	<	31		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:26	3878308
FS	Perfluorodecanoic acid (PFDoA)	L402	2.0	Filter Effluent	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:26	3878308
FS	Perfluorononanoic acid (PFNA)	L402	2.0	Filter Effluent	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:26	3878308
FS	Perfluorooctanesulfonic acid (PFOS)	L402	2.0	Filter Effluent	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:26	3878308
FS	N-ethyl Perfluorooctanesulfonamidoacetic acid	L402	2.0	Filter Effluent	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:26	3878308
FS	N-methyl Perfluorooctanesulfonamidoacetic acid	L402	2.0	Filter Effluent	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:26	3878308
FS	Perfluorooctanoic acid (PFOA)	L402	2.0	Filter Effluent	<	5.9		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:26	3878308
FS	Perfluorodecanoic acid (PFTDA)	L402	2.0	Filter Effluent	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:26	3878308
FS	Perfluoroundecanoic acid (PFUnA)	L402	2.0	Filter Effluent	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:26	3878308
FS	Perfluorododecanesulfonic acid (PFDS)	L402	2.0	Filter Effluent	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:26	3878308
FS	Perfluorododecanesulfonic acid (PFDS)	L402	2.0	Filter Effluent	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:26	3878308
FS	Perfluoroheptanesulfonic acid (PFHpS)	L402	2.0	Filter Effluent	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:26	3878308
FS	Perfluoroheptadecanoic acid (PFHxDA)	L402	2.0	Filter Effluent	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:26	3878308
FS	Perfluoro-2-methoxyethoxyacetic acid	L402	5.0	Filter Effluent	<	5.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:26	3878308
FS	Perfluoro-4-isopropoxybutanoic acid	L402	5.0	Filter Effluent	<	5.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:26	3878308
FS	Perfluoro-4-methoxybutanoic acid (PFMOBA)	L402	5.0	Filter Effluent	<	5.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:26	3878308
FS	Perfluoro-3-methoxypropanoic acid (PFMOPRA)	L402	5.0	Filter Effluent	<	5.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:26	3878308
FS	Perfluoronanesulfonic acid (PFNS)	L402	2.0	Filter Effluent	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:26	3878308
FS	Perfluorooctane sulfonamide (PFOSA)	L402	2.0	Filter Effluent	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:26	3878308
FS	Perfluoropentanoic acid (PFPeA)	L402	2.0	Filter Effluent	<	33		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:26	3878308
FS	Perfluoropentanesulfonic acid (PFPeS)	L402	2.0	Filter Effluent	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:26	3878308
FS	Perfluorotetradecanoic acid (PFTeDA)	L402	2.0	Filter Effluent	<	2.0		ng/L	---	---	---	---	0.9	03/09/2018 08:17	03/09/2018 03:26	3878308
FS	10:2 Fluorotelomer sulfonic acid (10:2 FTS)	L402	2.0	Biofiltration Filter #3	<	2.0		ng/L	---	---	---	---	0.92	03/09/2018 08:17	03/09/2018 03:52	3878309
FS	4:2 Fluorotelomer sulfonic acid (4:2 FTS)	L402	2.0	Biofiltration Filter #3	<	2.0		ng/L	---	---	---	---	0.92	03/09/2018 08:17	03/09/2018 03:52	3878309
FS	6:2 Fluorotelomer sulfonic acid (6:2 FTS)	L402	2.0	Biofiltration Filter #3	<	2.0		ng/L	---	---	---	---	0.92	03/09/2018 08:17	03/09/2018 03:52	3878309
FS	8:2 Fluorotelomer sulfonic acid (8:2 FTS)	L402	2.0	Biofiltration Filter #3	<	2.0		ng/L	---	---	---	---	0.92	03/09/2018 08:17	03/09/2018 03:52	3878309

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
FS	ADONA	L402	2.0	Biofiltration Filter #3	<	2.0		ng/L	---	---	---	---	0.92	03/08/2018 08:17	03/09/2018 03:52	3878309
FS	F-53B Major	L402	2.0	Biofiltration Filter #3	<	2.0		ng/L	---	---	---	---	0.92	03/08/2018 08:17	03/09/2018 03:52	3878309
FS	F-53B Minor	L402	2.0	Biofiltration Filter #3	<	2.0		ng/L	---	---	---	---	0.92	03/08/2018 08:17	03/09/2018 03:52	3878309
FS	GenX	L402	5.0	Biofiltration Filter #3	<	5.0		ng/L	---	---	---	---	0.92	03/08/2018 08:17	03/09/2018 03:52	3878309
FS	N-ethylperfluorooctane sulfonamide (NEFOSA)	L402	2.0	Biofiltration Filter #3	<	2.0		ng/L	---	---	---	---	0.92	03/08/2018 08:17	03/09/2018 03:52	3878309
FS	N-ethylperfluorooctane sulfonamide	L402	2.0	Biofiltration Filter #3	<	2.0		ng/L	---	---	---	---	0.92	03/08/2018 08:17	03/09/2018 03:52	3878309
FS	N-methylperfluorooctane sulfonamide	L402	2.0	Biofiltration Filter #3	<	2.0		ng/L	---	---	---	---	0.92	03/08/2018 08:17	03/09/2018 03:52	3878309
FS	N-methylperfluorooctane sulfonamideethanol	L402	2.0	Biofiltration Filter #3	<	2.0		ng/L	---	---	---	---	0.92	03/08/2018 08:17	03/09/2018 03:52	3878309
FS	N-methylperfluorooctane sulfonamideethanol	L402	2.0	Biofiltration Filter #3	<	2.0		ng/L	---	---	---	---	0.92	03/08/2018 08:17	03/09/2018 03:52	3878309
FS	Perfluorobutanesulfonic acid (PFBS)	L402	2.0	Biofiltration Filter #3	<	2.5		ng/L	---	---	---	---	0.92	03/08/2018 08:17	03/09/2018 03:52	3878309
FS	Perfluorobutanoic acid (PFBA)	L402	5.0	Biofiltration Filter #3	<	19		ng/L	---	---	---	---	0.92	03/08/2018 08:17	03/09/2018 03:52	3878309
FS	Perfluorodecanoic acid (PFDA)	L402	2.0	Biofiltration Filter #3	<	2.0		ng/L	---	---	---	---	0.92	03/08/2018 08:17	03/09/2018 03:52	3878309
FS	Perfluoroheptanoic acid (PFHpA)	L402	2.0	Biofiltration Filter #3	<	13		ng/L	---	---	---	---	0.92	03/08/2018 08:17	03/09/2018 03:52	3878309
FS	Perfluorohexanesulfonic acid (PFHxS)	L402	2.0	Biofiltration Filter #3	<	2.0		ng/L	---	---	---	---	0.92	03/08/2018 08:17	03/09/2018 03:52	3878309
FS	Perfluorohexanoic acid (PFHxA)	L402	2.0	Biofiltration Filter #3	<	31		ng/L	---	---	---	---	0.92	03/08/2018 08:17	03/09/2018 03:52	3878309
FS	Perfluorodecanoic acid (PFDoA)	L402	2.0	Biofiltration Filter #3	<	2.0		ng/L	---	---	---	---	0.92	03/08/2018 08:17	03/09/2018 03:52	3878309
FS	Perfluorononanoic acid (PFNA)	L402	2.0	Biofiltration Filter #3	<	2.0		ng/L	---	---	---	---	0.92	03/08/2018 08:17	03/09/2018 03:52	3878309
FS	Perfluorooctanesulfonic acid (PFOS)	L402	2.0	Biofiltration Filter #3	<	2.0		ng/L	---	---	---	---	0.92	03/08/2018 08:17	03/09/2018 03:52	3878309
FS	N-ethyl Perfluorooctanesulfonamidoacetic acid	L402	2.0	Biofiltration Filter #3	<	2.0		ng/L	---	---	---	---	0.92	03/08/2018 08:17	03/09/2018 03:52	3878309
FS	N-methyl Perfluorooctanesulfonamidoacetic acid	L402	2.0	Biofiltration Filter #3	<	2.0		ng/L	---	---	---	---	0.92	03/08/2018 08:17	03/09/2018 03:52	3878309
FS	Perfluorooctanoic acid (PFOA)	L402	2.0	Biofiltration Filter #3	<	5.2		ng/L	---	---	---	---	0.92	03/08/2018 08:17	03/09/2018 03:52	3878309
FS	Perfluorodecanoic acid (PFTDA)	L402	2.0	Biofiltration Filter #3	<	2.0		ng/L	---	---	---	---	0.92	03/08/2018 08:17	03/09/2018 03:52	3878309
FS	Perfluoroundecanoic acid (PFUnA)	L402	2.0	Biofiltration Filter #3	<	2.0		ng/L	---	---	---	---	0.92	03/08/2018 08:17	03/09/2018 03:52	3878309
FS	Perfluorododecanesulfonic acid (PFDS)	L402	2.0	Biofiltration Filter #3	<	2.0		ng/L	---	---	---	---	0.92	03/08/2018 08:17	03/09/2018 03:52	3878309
FS	Perfluorododecanoic acid (PFDoS)	L402	2.0	Biofiltration Filter #3	<	2.0		ng/L	---	---	---	---	0.92	03/08/2018 08:17	03/09/2018 03:52	3878309
FS	Perfluorodecanesulfonic acid (PFDS)	L402	2.0	Biofiltration Filter #3	<	2.0		ng/L	---	---	---	---	0.92	03/08/2018 08:17	03/09/2018 03:52	3878309
FS	Perfluoroheptanesulfonic acid (PFHpS)	L402	2.0	Biofiltration Filter #3	<	2.0		ng/L	---	---	---	---	0.92	03/08/2018 08:17	03/09/2018 03:52	3878309
FS	Perfluorohexadecanoic acid (PFHxDA)	L402	2.0	Biofiltration Filter #3	<	2.0		ng/L	---	---	---	---	0.92	03/08/2018 08:17	03/09/2018 03:52	3878309
FS	Perfluoro-2-methoxyethoxyacetic acid	L402	5.0	Biofiltration Filter #3	<	5.0		ng/L	---	---	---	---	0.92	03/08/2018 08:17	03/09/2018 03:52	3878309
FS	Perfluoro-4-isopropoxybutanoic acid	L402	5.0	Biofiltration Filter #3	<	5.0		ng/L	---	---	---	---	0.92	03/08/2018 08:17	03/09/2018 03:52	3878309
FS	Perfluoro-4-methoxybutanoic acid (PFMOBA)	L402	5.0	Biofiltration Filter #3	<	5.0		ng/L	---	---	---	---	0.92	03/08/2018 08:17	03/09/2018 03:52	3878309
FS	Perfluoro-3-methoxypropanoic acid (PFMOPRA)	L402	5.0	Biofiltration Filter #3	<	5.0		ng/L	---	---	---	---	0.92	03/08/2018 08:17	03/09/2018 03:52	3878309
FS	Perfluoronanesulfonic acid (PFNS)	L402	2.0	Biofiltration Filter #3	<	2.0		ng/L	---	---	---	---	0.92	03/08/2018 08:17	03/09/2018 03:52	3878309
FS	Perfluorooctane sulfonamide (PFOSA)	L402	2.0	Biofiltration Filter #3	<	2.0		ng/L	---	---	---	---	0.92	03/08/2018 08:17	03/09/2018 03:52	3878309
FS	Perfluoropentanoic acid (PFPeA)	L402	2.0	Biofiltration Filter #3	<	33		ng/L	---	---	---	---	0.92	03/08/2018 08:17	03/09/2018 03:52	3878309
FS	Perfluoropentanesulfonic acid (PFPeS)	L402	2.0	Biofiltration Filter #3	<	2.0		ng/L	---	---	---	---	0.92	03/08/2018 08:17	03/09/2018 03:52	3878309
FS	Perfluorotetradecanoic acid (PFTeDA)	L402	2.0	Biofiltration Filter #3	<	2.0		ng/L	---	---	---	---	0.92	03/08/2018 08:17	03/09/2018 03:52	3878309
FS	10:2 Fluorotelomer sulfonic acid (10:2 FTS)	L402	2.0	Biofiltration Filter #4	<	2.0		ng/L	---	---	---	---	0.91	03/08/2018 08:17	03/09/2018 04:18	3878310
FS	4:2 Fluorotelomer sulfonic acid (4:2 FTS)	L402	2.0	Biofiltration Filter #4	<	2.0		ng/L	---	---	---	---	0.91	03/08/2018 08:17	03/09/2018 04:18	3878310
FS	6:2 Fluorotelomer sulfonic acid (6:2 FTS)	L402	2.0	Biofiltration Filter #4	<	2.0		ng/L	---	---	---	---	0.91	03/08/2018 08:17	03/09/2018 04:18	3878310
FS	8:2 Fluorotelomer sulfonic acid (8:2 FTS)	L402	2.0	Biofiltration Filter #4	<	2.0		ng/L	---	---	---	---	0.91	03/08/2018 08:17	03/09/2018 04:18	3878310
FS	ADONA	L402	2.0	Biofiltration Filter #4	<	2.0		ng/L	---	---	---	---	0.91	03/08/2018 08:17	03/09/2018 04:18	3878310
FS	F-53B Major	L402	2.0	Biofiltration Filter #4	<	2.0		ng/L	---	---	---	---	0.91	03/08/2018 08:17	03/09/2018 04:18	3878310

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
FS	F-53B Minor	L402	2.0	Biofiltration Filter #4	<	2.0		ng/L	---	---	---	---	0.91	03/09/2018 08:17	03/09/2018 04:18	3878310
FS	GenX	L402	5.0	Biofiltration Filter #4	<	5.0		ng/L	---	---	---	---	0.91	03/09/2018 08:17	03/09/2018 04:18	3878310
FS	N-ethylperfluorooctane sulfonamide (NEFOSA)	L402	2.0	Biofiltration Filter #4	<	2.0		ng/L	---	---	---	---	0.91	03/09/2018 08:17	03/09/2018 04:18	3878310
FS	N-ethylperfluorooctane sulfonamideethanol	L402	2.0	Biofiltration Filter #4	<	2.0		ng/L	---	---	---	---	0.91	03/09/2018 08:17	03/09/2018 04:18	3878310
FS	N-methylperfluorooctane sulfonamide (NMeFOSA)	L402	2.0	Biofiltration Filter #4	<	2.0		ng/L	---	---	---	---	0.91	03/09/2018 08:17	03/09/2018 04:18	3878310
FS	N-methylperfluorooctane sulfonamideethanol	L402	2.0	Biofiltration Filter #4	<	2.0		ng/L	---	---	---	---	0.91	03/09/2018 08:17	03/09/2018 04:18	3878310
FS	Perfluorobutanesulfonic acid (PFBS)	L402	2.0	Biofiltration Filter #4	<	2.4		ng/L	---	---	---	---	0.91	03/09/2018 08:17	03/09/2018 04:18	3878310
FS	Perfluorobutanoic acid (PFBA)	L402	5.0	Biofiltration Filter #4	<	19		ng/L	---	---	---	---	0.91	03/09/2018 08:17	03/09/2018 04:18	3878310
FS	Perfluorodecanoic acid (PFDA)	L402	2.0	Biofiltration Filter #4	<	2.0		ng/L	---	---	---	---	0.91	03/09/2018 08:17	03/09/2018 04:18	3878310
FS	Perfluoroheptanoic acid (PFHpA)	L402	2.0	Biofiltration Filter #4	<	12		ng/L	---	---	---	---	0.91	03/09/2018 08:17	03/09/2018 04:18	3878310
FS	Perfluorohexanesulfonic acid (PFHxS)	L402	2.0	Biofiltration Filter #4	<	2.0		ng/L	---	---	---	---	0.91	03/09/2018 08:17	03/09/2018 04:18	3878310
FS	Perfluorohexanoic acid (PFHxA)	L402	2.0	Biofiltration Filter #4	<	30		ng/L	---	---	---	---	0.91	03/09/2018 08:17	03/09/2018 04:18	3878310
FS	Perfluorododecanoic acid (PFDoA)	L402	2.0	Biofiltration Filter #4	<	2.0		ng/L	---	---	---	---	0.91	03/09/2018 08:17	03/09/2018 04:18	3878310
FS	Perfluorononanoic acid (PFNA)	L402	2.0	Biofiltration Filter #4	<	2.0		ng/L	---	---	---	---	0.91	03/09/2018 08:17	03/09/2018 04:18	3878310
FS	Perfluorooctanesulfonic acid (PFOS)	L402	2.0	Biofiltration Filter #4	<	2.0		ng/L	---	---	---	---	0.91	03/09/2018 08:17	03/09/2018 04:18	3878310
FS	N-ethyl Perfluorooctanesulfonamidoacetic acid	L402	2.0	Biofiltration Filter #4	<	2.0		ng/L	---	---	---	---	0.91	03/09/2018 08:17	03/09/2018 04:18	3878310
FS	N-methyl Perfluorooctanesulfonamidoacetic acid	L402	2.0	Biofiltration Filter #4	<	2.0		ng/L	---	---	---	---	0.91	03/09/2018 08:17	03/09/2018 04:18	3878310
FS	Perfluorooctanoic acid (PFOA)	L402	2.0	Biofiltration Filter #4	<	4.3		ng/L	---	---	---	---	0.91	03/09/2018 08:17	03/09/2018 04:18	3878310
FS	Perfluorotridecanoic acid (PFTTDA)	L402	2.0	Biofiltration Filter #4	<	2.0		ng/L	---	---	---	---	0.91	03/09/2018 08:17	03/09/2018 04:18	3878310
FS	Perfluoroundecanoic acid (PFUnA)	L402	2.0	Biofiltration Filter #4	<	2.0		ng/L	---	---	---	---	0.91	03/09/2018 08:17	03/09/2018 04:18	3878310
FS	Perfluorododecanesulfonic acid (PFDS)	L402	2.0	Biofiltration Filter #4	<	2.0		ng/L	---	---	---	---	0.91	03/09/2018 08:17	03/09/2018 04:18	3878310
FS	Perfluorodecanesulfonic acid (PFDS)	L402	2.0	Biofiltration Filter #4	<	2.0		ng/L	---	---	---	---	0.91	03/09/2018 08:17	03/09/2018 04:18	3878310
FS	Perfluoroheptanesulfonic acid (PFHpS)	L402	2.0	Biofiltration Filter #4	<	2.0		ng/L	---	---	---	---	0.91	03/09/2018 08:17	03/09/2018 04:18	3878310
FS	Perfluorohexadecanoic acid (PFHxDA)	L402	2.0	Biofiltration Filter #4	<	2.0		ng/L	---	---	---	---	0.91	03/09/2018 08:17	03/09/2018 04:18	3878310
FS	Perfluoro-2-methoxyethoxyacetic acid	L402	5.0	Biofiltration Filter #4	<	5.0		ng/L	---	---	---	---	0.91	03/09/2018 08:17	03/09/2018 04:18	3878310
FS	Perfluoro-4-isopropoxybutanoic acid	L402	5.0	Biofiltration Filter #4	<	5.0		ng/L	---	---	---	---	0.91	03/09/2018 08:17	03/09/2018 04:18	3878310
FS	Perfluoro-4-methoxybutanoic acid (PFMOBA)	L402	5.0	Biofiltration Filter #4	<	5.0		ng/L	---	---	---	---	0.91	03/09/2018 08:17	03/09/2018 04:18	3878310
FS	Perfluoro-3-methoxypropanoic acid (PFMOPRA)	L402	5.0	Biofiltration Filter #4	<	5.0		ng/L	---	---	---	---	0.91	03/09/2018 08:17	03/09/2018 04:18	3878310
FS	Perfluoronanesulfonic acid (PFNS)	L402	2.0	Biofiltration Filter #4	<	2.0		ng/L	---	---	---	---	0.91	03/09/2018 08:17	03/09/2018 04:18	3878310
FS	Perfluorooctane sulfonamide (PFOSA)	L402	2.0	Biofiltration Filter #4	<	2.0		ng/L	---	---	---	---	0.91	03/09/2018 08:17	03/09/2018 04:18	3878310
FS	Perfluoropentanoic acid (PFPeA)	L402	2.0	Biofiltration Filter #4	<	32		ng/L	---	---	---	---	0.91	03/09/2018 08:17	03/09/2018 04:18	3878310
FS	Perfluoropentanesulfonic acid (PFPeS)	L402	2.0	Biofiltration Filter #4	<	2.0		ng/L	---	---	---	---	0.91	03/09/2018 08:17	03/09/2018 04:18	3878310
FS	Perfluorotetradecanoic acid (PFTeDA)	L402	2.0	Biofiltration Filter #4	<	2.0		ng/L	---	---	---	---	0.91	03/09/2018 08:17	03/09/2018 04:18	3878310
FS	10:2 Fluorotelomer sulfonic acid (10:2 FTS)	L402	2.0	Biofiltration Filter #5	<	2.0		ng/L	---	---	---	---	0.89	03/09/2018 08:17	03/09/2018 04:44	3878311
FS	4:2 Fluorotelomer sulfonic acid (4:2 FTS)	L402	2.0	Biofiltration Filter #5	<	2.0		ng/L	---	---	---	---	0.89	03/09/2018 08:17	03/09/2018 04:44	3878311
FS	6:2 Fluorotelomer sulfonic acid (6:2 FTS)	L402	2.0	Biofiltration Filter #5	<	2.0		ng/L	---	---	---	---	0.89	03/09/2018 08:17	03/09/2018 04:44	3878311
FS	8:2 Fluorotelomer sulfonic acid (8:2 FTS)	L402	2.0	Biofiltration Filter #5	<	2.0		ng/L	---	---	---	---	0.89	03/09/2018 08:17	03/09/2018 04:44	3878311
FS	ADONA	L402	2.0	Biofiltration Filter #5	<	2.0		ng/L	---	---	---	---	0.89	03/09/2018 08:17	03/09/2018 04:44	3878311
FS	F-53B Major	L402	2.0	Biofiltration Filter #5	<	2.0		ng/L	---	---	---	---	0.89	03/09/2018 08:17	03/09/2018 04:44	3878311
FS	F-53B Minor	L402	2.0	Biofiltration Filter #5	<	2.0		ng/L	---	---	---	---	0.89	03/09/2018 08:17	03/09/2018 04:44	3878311
FS	GenX	L402	5.0	Biofiltration Filter #5	<	5.0		ng/L	---	---	---	---	0.89	03/09/2018 08:17	03/09/2018 04:44	3878311

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
FS	N-ethylperfluorooctane sulfonamide (NEFOSA)	L402	2.0	Biofiltration Filter #5	<	2.0		ng/L	---	---	---	---	0.89	03/09/2018 08:17	03/09/2018 04:44	3878311
FS	N-ethylperfluorooctane sulfonamidoethanol	L402	2.0	Biofiltration Filter #5	<	2.0		ng/L	---	---	---	---	0.89	03/09/2018 08:17	03/09/2018 04:44	3878311
FS	N-methylperfluorooctane sulfonamide (NMeFOSA)	L402	2.0	Biofiltration Filter #5	<	2.0		ng/L	---	---	---	---	0.89	03/09/2018 08:17	03/09/2018 04:44	3878311
FS	N-methylperfluorooctane sulfonamidoethanol	L402	2.0	Biofiltration Filter #5	<	2.0		ng/L	---	---	---	---	0.89	03/09/2018 08:17	03/09/2018 04:44	3878311
FS	Perfluorobutanesulfonic acid (PFBS)	L402	2.0	Biofiltration Filter #5	<	2.9		ng/L	---	---	---	---	0.89	03/09/2018 08:17	03/09/2018 04:44	3878311
FS	Perfluorobutanoic acid (PFBA)	L402	5.0	Biofiltration Filter #5	<	2.0		ng/L	---	---	---	---	0.89	03/09/2018 08:17	03/09/2018 04:44	3878311
FS	Perfluorodecanoic acid (PFDA)	L402	2.0	Biofiltration Filter #5	<	2.0		ng/L	---	---	---	---	0.89	03/09/2018 08:17	03/09/2018 04:44	3878311
FS	Perfluoroheptanoic acid (PFHpA)	L402	2.0	Biofiltration Filter #5	<	16		ng/L	---	---	---	---	0.89	03/09/2018 08:17	03/09/2018 04:44	3878311
FS	Perfluorohexanesulfonic acid (PFHxS)	L402	2.0	Biofiltration Filter #5	<	2.0		ng/L	---	---	---	---	0.89	03/09/2018 08:17	03/09/2018 04:44	3878311
FS	Perfluorohexanoic acid (PFHxA)	L402	2.0	Biofiltration Filter #5	<	34		ng/L	---	---	---	---	0.89	03/09/2018 08:17	03/09/2018 04:44	3878311
FS	Perfluorododecanoic acid (PFDoA)	L402	2.0	Biofiltration Filter #5	<	2.0		ng/L	---	---	---	---	0.89	03/09/2018 08:17	03/09/2018 04:44	3878311
FS	Perfluorononanoic acid (PFNA)	L402	2.0	Biofiltration Filter #5	<	2.0		ng/L	---	---	---	---	0.89	03/09/2018 08:17	03/09/2018 04:44	3878311
FS	Perfluorooctanesulfonic acid (PFOS)	L402	2.0	Biofiltration Filter #5	<	2.0		ng/L	---	---	---	---	0.89	03/09/2018 08:17	03/09/2018 04:44	3878311
FS	N-ethyl Perfluorooctanesulfonamidoacetic acid	L402	2.0	Biofiltration Filter #5	<	2.0		ng/L	---	---	---	---	0.89	03/09/2018 08:17	03/09/2018 04:44	3878311
FS	N-methyl Perfluorooctanesulfonamidoacetic acid	L402	2.0	Biofiltration Filter #5	<	2.0		ng/L	---	---	---	---	0.89	03/09/2018 08:17	03/09/2018 04:44	3878311
FS	Perfluorooctanoic acid (PFOA)	L402	2.0	Biofiltration Filter #5	<	6.5		ng/L	---	---	---	---	0.89	03/09/2018 08:17	03/09/2018 04:44	3878311
FS	Perfluorodecanoic acid (PFTrDA)	L402	2.0	Biofiltration Filter #5	<	2.0		ng/L	---	---	---	---	0.89	03/09/2018 08:17	03/09/2018 04:44	3878311
FS	Perfluoroundecanoic acid (PFUnA)	L402	2.0	Biofiltration Filter #5	<	2.0		ng/L	---	---	---	---	0.89	03/09/2018 08:17	03/09/2018 04:44	3878311
FS	Perfluorododecanesulfonic acid (PFDS)	L402	2.0	Biofiltration Filter #5	<	2.0		ng/L	---	---	---	---	0.89	03/09/2018 08:17	03/09/2018 04:44	3878311
FS	Perfluorodecanesulfonic acid (PFDS)	L402	2.0	Biofiltration Filter #5	<	2.0		ng/L	---	---	---	---	0.89	03/09/2018 08:17	03/09/2018 04:44	3878311
FS	Perfluoroheptanesulfonic acid (PFHpS)	L402	2.0	Biofiltration Filter #5	<	2.0		ng/L	---	---	---	---	0.89	03/09/2018 08:17	03/09/2018 04:44	3878311
FS	Perfluorohexadecanoic acid (PFHxDA)	L402	2.0	Biofiltration Filter #5	<	2.0		ng/L	---	---	---	---	0.89	03/09/2018 08:17	03/09/2018 04:44	3878311
FS	Perfluoro-2-methoxyethoxyacetic acid	L402	5.0	Biofiltration Filter #5	<	5.0		ng/L	---	---	---	---	0.89	03/09/2018 08:17	03/09/2018 04:44	3878311
FS	Perfluoro-4-isopropoxybutanoic acid	L402	5.0	Biofiltration Filter #5	<	5.0		ng/L	---	---	---	---	0.89	03/09/2018 08:17	03/09/2018 04:44	3878311
FS	Perfluoro-4-methoxybutanoic acid (PFMOBA)	L402	5.0	Biofiltration Filter #5	<	5.0		ng/L	---	---	---	---	0.89	03/09/2018 08:17	03/09/2018 04:44	3878311
FS	Perfluoro-3-methoxypropanoic acid (PFMOPRA)	L402	5.0	Biofiltration Filter #5	<	5.0		ng/L	---	---	---	---	0.89	03/09/2018 08:17	03/09/2018 04:44	3878311
FS	Perfluoronanesulfonic acid (PFNS)	L402	2.0	Biofiltration Filter #5	<	2.0		ng/L	---	---	---	---	0.89	03/09/2018 08:17	03/09/2018 04:44	3878311
FS	Perfluorooctane sulfonamide (PFOSA)	L402	2.0	Biofiltration Filter #5	<	2.0		ng/L	---	---	---	---	0.89	03/09/2018 08:17	03/09/2018 04:44	3878311
FS	Perfluoropentanoic acid (PFPeA)	L402	2.0	Biofiltration Filter #5	<	34		ng/L	---	---	---	---	0.89	03/09/2018 08:17	03/09/2018 04:44	3878311
FS	Perfluoropentanesulfonic acid (PFPeS)	L402	2.0	Biofiltration Filter #5	<	2.0		ng/L	---	---	---	---	0.89	03/09/2018 08:17	03/09/2018 04:44	3878311
FS	Perfluorotetradecanoic acid (PFTeDA)	L402	2.0	Biofiltration Filter #5	<	2.0		ng/L	---	---	---	---	0.89	03/09/2018 08:17	03/09/2018 04:44	3878311
CCH	IS-4-2-FTS-13C2	L402	N/A	---		217468	520489	ng/L	42	50 - 150	---	---	1.0	03/09/2018 14:29	03/09/2018 06:54	3886493
CCH	IS-6-2-FTS-13C2	L402	N/A	---		29221	82370	ng/L	35	50 - 150	---	---	1.0	03/09/2018 14:29	03/09/2018 06:54	3886493
CCH	IS-8-2-FTS-13C2	L402	N/A	---		139756	400513	ng/L	35	50 - 150	---	---	1.0	03/09/2018 14:29	03/09/2018 06:54	3886493
CCH	IS-GenX-13C3	L402	N/A	---		4871	6129	ng/L	79	50 - 150	---	---	1.0	03/09/2018 14:29	03/09/2018 06:54	3886493
CCH	IS-NEIFOSAA-d5	L402	N/A	---		223292	552142	ng/L	40	50 - 150	---	---	1.0	03/09/2018 14:29	03/09/2018 06:54	3886493
CCH	IS-NEIFOSE-d9	L402	N/A	---		57396	96472	ng/L	59	50 - 150	---	---	1.0	03/09/2018 14:29	03/09/2018 06:54	3886493
CCH	IS-NIMEFOSAA-d3	L402	N/A	---		289426	696176	ng/L	42	50 - 150	---	---	1.0	03/09/2018 14:29	03/09/2018 06:54	3886493
CCH	IS-NIMEFOSA-d3	L402	N/A	---		302364	588018	ng/L	51	50 - 150	---	---	1.0	03/09/2018 14:29	03/09/2018 06:54	3886493
CCH	IS-NIMEFOSE-d7	L402	N/A	---		56978	100564	ng/L	57	50 - 150	---	---	1.0	03/09/2018 14:29	03/09/2018 06:54	3886493
CCH	IS-PFBA-13C4	L402	N/A	---		1102910	1971023	ng/L	56	50 - 150	---	---	1.0	03/09/2018 14:29	03/09/2018 06:54	3886493

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCH	IS-PFBS-13C3	L402	N/A	---		286491	497347	ng/L	58	50 - 150	---	---	1.0	03/09/2018 14:29	03/09/2018 06:54	3886493
CCH	IS-PFDA-13C2	L402	N/A	---		1396140	3124910	ng/L	45	50 - 150	---	---	1.0	03/08/2018 14:29	03/09/2018 06:54	3886493
CCH	IS-PFDoA-13C2	L402	N/A	---		1063870	2539471	ng/L	42	50 - 150	---	---	1.0	03/08/2018 14:29	03/09/2018 06:54	3886493
CCH	IS-PFHpa-13C4	L402	N/A	---		1974050	3493018	ng/L	57	50 - 150	---	---	1.0	03/08/2018 14:29	03/09/2018 06:54	3886493
CCH	IS-PFHXA-13C2	L402	N/A	---		2448900	4133503	ng/L	59	50 - 150	---	---	1.0	03/08/2018 14:29	03/09/2018 06:54	3886493
CCH	IS-PFOsa-13C8	L402	N/A	---		486629	1118883	ng/L	43	50 - 150	---	---	1.0	03/08/2018 14:29	03/09/2018 06:54	3886493
CCH	IS-NEIFOSA-d5	L402	N/A	---		261639	524732	ng/L	50	50 - 150	---	---	1.0	03/08/2018 14:29	03/09/2018 06:54	3886493
CCH	IS-PFHXS-13C3	L402	N/A	---		307833	498313	ng/L	62	50 - 150	---	---	1.0	03/08/2018 14:29	03/09/2018 06:54	3886493
CCH	IS-PFNA-13C9	L402	N/A	---		1650400	3311238	ng/L	50	50 - 150	---	---	1.0	03/08/2018 14:29	03/09/2018 06:54	3886493
CCH	IS-PFOA-13C2	L402	N/A	---		1947810	3737396	ng/L	52	50 - 150	---	---	1.0	03/08/2018 14:29	03/09/2018 06:54	3886493
CCH	IS-PFOS-13C4	L402	N/A	---		226912	443473	ng/L	51	50 - 150	---	---	1.0	03/08/2018 14:29	03/09/2018 06:54	3886493
CCH	IS-PFPeA-13C5	L402	N/A	---		779924	1366858	ng/L	57	50 - 150	---	---	1.0	03/08/2018 14:29	03/09/2018 06:54	3886493
CCH	IS-PFTeDA-13C2	L402	N/A	---		768112	1588255	ng/L	48	50 - 150	---	---	1.0	03/08/2018 14:29	03/09/2018 06:54	3886493
CCH	IS-PFUnA-13C7	L402	N/A	---		953195	2529015	ng/L	38	50 - 150	---	---	1.0	03/08/2018 14:29	03/09/2018 06:54	3886493
CCH	10:2 Fluorotelomer sulfonic acid (10:2 FTS)	L402	2.0	---		149.3730	200	ng/L	75	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 06:54	3886493
CCH	4:2 Fluorotelomer sulfonic acid (4:2 FTS)	L402	2.0	---		199.2770	200	ng/L	100	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 06:54	3886493
CCH	6:2 Fluorotelomer sulfonic acid (6:2 FTS)	L402	2.0	---		203.1610	200	ng/L	102	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 06:54	3886493
CCH	8:2 Fluorotelomer sulfonic acid (8:2 FTS)	L402	2.0	---		195.1380	200	ng/L	98	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 06:54	3886493
CCH	ADONA	L402	2.0	---		206.0480	200	ng/L	103	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 06:54	3886493
CCH	F-53B Major	L402	2.0	---		211.6290	200	ng/L	106	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 06:54	3886493
CCH	F-53B Minor	L402	2.0	---		236.7900	200	ng/L	118	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 06:54	3886493
CCH	GenX	L402	5.0	---		479.6150	500	ng/L	96	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 06:54	3886493
CCH	N-ethylperfluorooctane sulfonamide (NEFOSA)	L402	2.0	---		198.8930	200	ng/L	99	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 06:54	3886493
CCH	N-ethylperfluorooctane sulfonamidoethanol	L402	2.0	---		184.6290	200	ng/L	92	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 06:54	3886493
CCH	N-methylperfluorooctane sulfonamide (NMeFOSA)	L402	2.0	---		198.1680	200	ng/L	99	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 06:54	3886493
CCH	N-methylperfluorooctane sulfonamidoethanol	L402	2.0	---		191.7090	200	ng/L	96	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 06:54	3886493
CCH	Perfluorobutanesulfonic acid (PFBS)	L402	2.0	---		200.3130	200	ng/L	100	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 06:54	3886493
CCH	Perfluorobutanoic acid (PFBA)	L402	5.0	---		501.2820	500	ng/L	100	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 06:54	3886493
CCH	Perfluorodecanoic acid (PFDA)	L402	2.0	---		201.8280	200	ng/L	101	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 06:54	3886493
CCH	Perfluoroheptanoic acid (PFHpA)	L402	2.0	---		197.9100	200	ng/L	99	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 06:54	3886493
CCH	Perfluorohexanesulfonic acid (PFHXS)	L402	2.0	---		200.6550	200	ng/L	100	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 06:54	3886493
CCH	Perfluorohexanoic acid (PFHxA)	L402	2.0	---		201.3450	200	ng/L	101	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 06:54	3886493
CCH	Perfluorodecanoic acid (PFDoA)	L402	2.0	---		200.9630	200	ng/L	100	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 06:54	3886493
CCH	Perfluorononanoic acid (PFNA)	L402	2.0	---		198.6910	200	ng/L	99	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 06:54	3886493
CCH	Perfluorooctanesulfonic acid (PFOS)	L402	2.0	---		203.1710	200	ng/L	102	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 06:54	3886493
CCH	N-ethyl Perfluorooctanesulfonamidoacetic acid	L402	2.0	---		202.5320	200	ng/L	101	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 06:54	3886493
CCH	N-methyl Perfluorooctanesulfonamidoacetic acid	L402	2.0	---		204.4850	200	ng/L	102	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 06:54	3886493
CCH	Perfluorooctanoic acid (PFOA)	L402	2.0	---		197.8490	200	ng/L	99	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 06:54	3886493
CCH	Perfluorotridecanoic acid (PFTrDA)	L402	2.0	---		180.8370	200	ng/L	90	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 06:54	3886493
CCH	Perfluoroundecanoic acid (PFUnA)	L402	2.0	---		200.4440	200	ng/L	100	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 06:54	3886493
CCH	Perfluorododecanesulfonic acid (PFDoS)	L402	2.0	---		175.2410	200	ng/L	88	70 - 130	---	---	1.0	03/08/2018 14:29	03/09/2018 06:54	3886493

QC Summary Report (cont.)

Sample Type	Analyte	Method	MRL	Client ID	Result Flag	Amount	Target	Units	% Recovery	Recovery Limits	RPD	RPD Limit	Dil Factor	Extracted	Analyzed	EEA ID #
CCH	Perfluorodecanesulfonic acid (PFDS)	L402	2.0	---		243.8160	200	ng/L	122	70 - 130	---	---	1.0	03/09/2018 14:29	03/09/2018 06:54	3886493
CCH	Perfluorohexanesulfonic acid (PFH6S)	L402	2.0	---		191.1710	200	ng/L	96	70 - 130	---	---	1.0	03/09/2018 14:29	03/09/2018 06:54	3886493
CCH	Perfluorohexadecanoic acid (PFH6DA)	L402	2.0	---		202.7000	200	ng/L	101	70 - 130	---	---	1.0	03/09/2018 14:29	03/09/2018 06:54	3886493
CCH	Perfluoro-2-methoxyethoxyacetic acid	L402	5.0	---		544.0620	500	ng/L	109	70 - 130	---	---	1.0	03/09/2018 14:29	03/09/2018 06:54	3886493
CCH	Perfluoro-4-isopropoxybutanoic acid	L402	5.0	---		522.9290	500	ng/L	105	70 - 130	---	---	1.0	03/09/2018 14:29	03/09/2018 06:54	3886493
CCH	Perfluoro-4-methoxybutanoic acid (PFMOBA)	L402	5.0	---		516.8980	500	ng/L	103	70 - 130	---	---	1.0	03/09/2018 14:29	03/09/2018 06:54	3886493
CCH	Perfluoro-3-methoxypropanoic acid (PFMOPRA)	L402	5.0	---		546.4320	500	ng/L	109	70 - 130	---	---	1.0	03/09/2018 14:29	03/09/2018 06:54	3886493
CCH	Perfluorononanesulfonic acid (PFNS)	L402	2.0	---		204.2790	200	ng/L	102	70 - 130	---	---	1.0	03/09/2018 14:29	03/09/2018 06:54	3886493
CCH	Perfluorooctane sulfonamide (PFOSA)	L402	2.0	---		196.9650	200	ng/L	98	70 - 130	---	---	1.0	03/09/2018 14:29	03/09/2018 06:54	3886493
CCH	Perfluoropentanoic acid (PFPeA)	L402	2.0	---		200.3470	200	ng/L	100	70 - 130	---	---	1.0	03/09/2018 14:29	03/09/2018 06:54	3886493
CCH	Perfluoropentanesulfonic acid (PFPeS)	L402	2.0	---		214.5990	200	ng/L	107	70 - 130	---	---	1.0	03/09/2018 14:29	03/09/2018 06:54	3886493
CCH	Perfluorotetradecanoic acid (PFTeDA)	L402	2.0	---		201.5410	200	ng/L	101	70 - 130	---	---	1.0	03/09/2018 14:29	03/09/2018 06:54	3886493

Sample Type Key

<u>Type (Abbr.)</u>	<u>Sample Type</u>	<u>Type (Abbr.)</u>	<u>Sample Type</u>
CCH	Continuing Calibration High		
CCL	Continuing Calibration Low		
CCM	Continuing Calibration Mid		
FS	Field Sample		
FBL	Fortified Blank Low		
LRB	Laboratory Reagent Blank		

END OF REPORT